

REF		Σ	SYSTEM
11776193122*	11776193500	100	cobas e 411
11776193214*			cobas e 601 cobas e 602

* Some kits shown may not be available in all countries.

English

System information

For **cobas e 411** analyzer: test number 351

For **cobas e 601** and **cobas e 602** analyzers: Application Code Number 054

Please note

The measured CA 19-9 value of a patient's sample can vary depending on the testing procedure used. The laboratory finding must therefore always contain a statement on the CA 19-9 assay method used. CA 19-9 values determined on patient samples by different testing procedures cannot be directly compared with one another and could be the cause of erroneous medical interpretations. If there is a change in the CA 19-9 assay procedure used while monitoring therapy, then the CA 19-9 values obtained upon changing over to the new procedure must be confirmed by parallel measurements with both methods.

Intended use

Immunoassay for the in vitro quantitative determination of CA 19-9 in human serum and plasma.

The electrochemiluminescence immunoassay "ECLIA" is intended for use on Elecsys and **cobas e** immunoassay analyzers.

Summary

CA 19-9 (carbohydrate antigen 19-9 or sialylated Lewis (a) antigen) is a biomarker which is primarily used in the management of pancreatic cancer patients in addition to other diagnostic methods.¹ The CA 19-9 antibody binds to the Lewis (a) antigen on a mucin.^{2,3} Elevated concentrations are frequently present in the blood of patients with various gastrointestinal conditions, such as pancreatic-, colorectal-, gastric-, hepatocellular- and cholangiocellular carcinomas.⁴

No data exist today which support the use of CA 19-9 in screening for malignancies⁵ also concerning the fact that approximately 6 % of the population belong to the Lewis (a-/b-) blood group, lacking the antigenic determinant CA 19-9 and will therefore not release CA 19-9 even when a malignancy is present. This must be taken into account when interpreting the findings.⁶

Among non-malignant conditions, obstructive jaundice is frequently associated with increases in CA 19-9⁷ and unspecific elevation of CA 19-9 in serum reflects both inflammatory hypersecretion and leakage of biliary mucins into serum.⁸ CA 19-9 levels have also been reported in benign diseases like cystic fibrosis, hydronephrosis, and Hashimoto's thyroiditis.⁹

In addition, there is a strong correlation between the serum CA 19-9 concentration and the degree of cholestasis as well as the levels of alkaline phosphatase and bilirubin during acute liver failure, acute hepatitis or chronic liver diseases.^{10,11} The common underlying mechanism for elevations in non-malignant conditions is probably inflammatory hypersecretion of CA 19-9 by epithelial cells.

In pancreatic cancer, levels > 100 U/mL are highly suggestive of unresectability or metastatic disease and levels < 100 U/mL imply a likely resectable disease.¹²

The European Group of Tumor Markers (EGTM) advise that CA 19-9 may be used as a diagnostic aid and for monitoring therapy in patients with pancreatic adenocarcinoma.¹³ CA 19-9 has been found to be prognostic for survival following resection of pancreatic ductal adenocarcinoma.¹⁴

In hepatobiliary carcinoma, CA 19-9 independently predicted a 2.6-fold increased mortality in a prospectively collected group of HCC patients in a multivariable analysis.¹⁵ In colorectal cancer, CA 19-9 is described as an additional marker for disease monitoring in patients without an increase in CEA.¹⁶

Test principle

Sandwich principle. Total duration of assay: 18 minutes.



100

SYSTEM

cobas e 411
cobas e 601
cobas e 602

- 1st incubation: 10 µL of sample, a biotinylated monoclonal CA 19-9-specific antibody, and a monoclonal CA 19-9-specific antibody labeled with a ruthenium complex^{a)} form a sandwich complex.
- 2nd incubation: After addition of streptavidin-coated microparticles, the complex becomes bound to the solid phase via interaction of biotin and streptavidin.
- The reaction mixture is aspirated into the measuring cell where the microparticles are magnetically captured onto the surface of the electrode. Unbound substances are then removed with ProCell/ProCell M. Application of a voltage to the electrode then induces chemiluminescent emission which is measured by a photomultiplier.
- Results are determined via a calibration curve which is instrument-specifically generated by 2-point calibration and a master curve provided via the reagent barcode or e-barcode.

a) Tris(2,2'-bipyridyl)ruthenium(II)-complex (Ru(bpy)₃²⁺)

Reagents - working solutions

The reagent rackpack is labeled as CA19-9.

- M Streptavidin-coated microparticles (transparent cap), 1 bottle, 6.5 mL:
Streptavidin-coated microparticles 0.72 mg/mL; preservative.
- R1 Anti-CA 19-9-Ab~biotin (gray cap), 1 bottle, 10 mL:
Biotinylated monoclonal anti-CA 19-9 antibody (mouse) 3 mg/L,
phosphate buffer 100 mmol/L, pH 6.5; preservative.
- R2 Anti-CA 19-9-Ab~Ru(bpy)₃²⁺ (black cap), 1 bottle, 10 mL:
Monoclonal anti-CA 19-9 antibody (mouse) labeled with ruthenium
complex 4 mg/L; phosphate buffer 100 mmol/L, pH 6.5; preservative.

Precautions and warnings

For in vitro diagnostic use for health care professionals. Exercise the normal precautions required for handling all laboratory reagents.

Infectious or microbial waste:

Warning: handle waste as potentially biohazardous material. Dispose of waste according to accepted laboratory instructions and procedures.

Environmental hazards:

Apply all relevant local disposal regulations to determine the safe disposal.

Safety data sheet available for professional user on request.

This kit contains components classified as follows in accordance with the Regulation (EC) No. 1272/2008:



Warning

H317 May cause an allergic skin reaction.

Prevention:

P261 Avoid breathing dust/fume/gas/mist/vapours/spray.

P272 Contaminated work clothing should not be allowed out of the workplace.

P280 Wear protective gloves.

Response:

P333 + P313 If skin irritation or rash occurs: Get medical advice/attention.

P362 + P364 Take off contaminated clothing and wash it before reuse.

Disposal:

P501 Dispose of contents/container to an approved waste disposal plant.

Product safety labeling follows EU GHS guidance.

Contact phone: all countries: +49-621-7590

Avoid foam formation in all reagents and sample types (specimens, calibrators and controls).

Reagent handling

The reagents in the kit have been assembled into a ready-for-use unit that cannot be separated.

All information required for correct operation is read in from the respective reagent barcodes.

Storage and stability

Store at 2-8 °C.

Do not freeze.

Store the Elecsys reagent kit **upright** in order to ensure complete availability of the microparticles during automatic mixing prior to use.

Stability:	
unopened at 2-8 °C	up to the stated expiration date
after opening at 2-8 °C	8 weeks
on cobas e 601 and cobas e 602	6 weeks
on cobas e 411	8 weeks

Specimen collection and preparation

Only the specimens listed below were tested and found acceptable.

Serum collected using standard sampling tubes or tubes containing separating gel.

Li-heparin, K₂-EDTA and K₃-EDTA plasma.

Do not use sodium citrate plasma.

Criterion: Slope 0.9-1.1 + coefficient of correlation ≥ 0.95.

Stable for 14 days at 2-8 °C, 5 days at 20-25 °C, 3 months at -20 °C (± 5 °C). Freeze only once.

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube manufacturer.

Centrifuge samples containing precipitates before performing the assay.

Do not use heat-inactivated samples.

Do not use samples and controls stabilized with azide.

Ensure the samples, calibrators and controls are at 20-25 °C prior to measurement.

Due to possible evaporation effects, samples, calibrators and controls on the analyzers should be analyzed/measured within 2 hours.

Materials provided

See "Reagents – working solutions" section for reagents.

Materials required (but not provided)

- [REF] 11776215122, CA 19-9 CalSet, for 4 x 1.0 mL
- [REF] 11776452122, PreciControl Tumor Marker, for 4 x 3.0 mL
- [REF] 11732277122, Diluent Universal, 2 x 16 mL sample diluent or [REF] 03183971122, Diluent Universal, 2 x 36 mL sample diluent
- General laboratory equipment
- **cobas e** analyzer

Additional materials for the **cobas e 411** analyzer:

- [REF] 11662988122, ProCell, 6 x 380 mL system buffer

- [REF] 11662970122, CleanCell, 6 x 380 mL measuring cell cleaning solution
- [REF] 11930346122, Elecsys SysWash, 1 x 500 mL washwater additive
- [REF] 11933159001, Adapter for SysClean
- [REF] 11706802001, AssayCup, 60 x 60 reaction cups
- [REF] 11706799001, AssayTip, 30 x 120 pipette tips
- [REF] 11800507001, Clean-Liner

Additional materials for **cobas e 601** and **cobas e 602** analyzers:

- [REF] 04880340190, ProCell M, 2 x 2 L system buffer
- [REF] 04880293190, CleanCell M, 2 x 2 L measuring cell cleaning solution
- [REF] 03023141001, PC/CC-Cups, 12 cups to prewarm ProCell M and CleanCell M before use
- [REF] 03005712190, ProbeWash M, 12 x 70 mL cleaning solution for run finalization and rinsing during reagent change
- [REF] 12102137001, AssayTip/AssayCup, 48 magazines x 84 reaction cups or pipette tips, waste bags
- [REF] 03023150001, WasteLiner, waste bags
- [REF] 03027651001, SysClean Adapter M

Additional materials for all analyzers:

- [REF] 11298500316, ISE Cleaning Solution/Elecsys SysClean, 5 x 100 mL system cleaning solution

Assay

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

Resuspension of the microparticles takes place automatically prior to use. Read in the test-specific parameters via the reagent barcode. If in exceptional cases the barcode cannot be read, enter the 15-digit sequence of numbers.

Bring the cooled reagents to approximately 20 °C and place on the reagent disk (20 °C) of the analyzer. Avoid foam formation. The system automatically regulates the temperature of the reagents and the opening/closing of the bottles.

Calibration

Traceability: This method has been standardized against the Enzymun-Test CA 19-9 method.

Every Elecsys reagent set has a barcoded label containing specific information for calibration of the particular reagent lot. The predefined master curve is adapted to the analyzer using the relevant CalSet.

Calibration frequency: Calibration must be performed once per reagent lot using fresh reagent (i.e. not more than 24 hours since the reagent kit was registered on the analyzer).

Calibration interval may be extended based on acceptable verification of calibration by the laboratory.

Renewed calibration is recommended as follows:

- after 12 weeks when using the same reagent lot
- after 7 days (when using the same reagent kit on the analyzer)
- as required: e.g. quality control findings outside the defined limits

Quality control

For quality control, use PreciControl Tumor Marker.

In addition, other suitable control material can be used.

Controls for the various concentration ranges should be run individually at least once every 24 hours when the test is in use, once per reagent kit, and following each calibration.

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

If necessary, repeat the measurement of the samples concerned.

Follow the applicable government regulations and local guidelines for quality control.

- 9 Parra JL, Kaplan S, Barkin JS, et al. Elevated CA 19-9 Caused by Hashimoto's Thyroiditis: Review of the Benign Causes of Increased CA 19-9 Level. *Dig Dis Sci* 2005;50(4):694-695.
- 10 Maestranzi S, Przemioslo R, Mitchell H, et al. The effect of benign and malignant liver disease on the tumour markers CA19.9 and CEA. *Ann Clin Biochem* 1998;35(1):99-103.
- 11 Halme L, Karkkainen P, Isoniemi H, et al. Carbohydrate 19-9 antigen as a marker of non-malignant hepatocytic ductular transformation in patients with acute liver failure. A comparison with alpha-fetoprotein and carcinoembryonic antigen. *Scand J Gastroenterol* 1999;34(4):426-431.
- 12 Ballehaninna UK, Chamberlain RS. The clinical utility of serum CA 19-9 in the diagnosis, prognosis and management of pancreatic adenocarcinoma: An evidence based appraisal. *Journal of Gastrointestinal Oncology* 2012;3(2):105-119.
- 13 Duffy MJ, Sturgeon C, Lamerz R, et al. Tumor markers in pancreatic cancer: a European Group on Tumor Markers (EGTM) status report. *Annals of Oncol* 2010;21: 441-447.
- 14 NCCN Guidelines 2.2015 Pancreatic Adenocarcinoma; http://www.nccn.org/professionals/physician_gls/pdf/pancreatic.pdf
- 15 Hsu CC, Goyal A, Iuga A, et al. Elevated CA19-9 Is Associated With Increased Mortality In A Prospective Cohort Of Hepatocellular Carcinoma Patients. *Clin Transl Gastroenterol* 2015; 6: e74.
- 16 Stikma J, Grootendorst DC, van den Linden PW. CA19-9 As a Marker in Addition to CEA to Monitor Colorectal Cancer. *Clin Colorectal Cancer* 2014;13(4): 239-244.
- 17 Suresh MR. Immunoassays for cancer-associated carbohydrate antigens. *Semin Cancer Biol* 1991 Dec;2(6):367-377.
- 18 Bablok W, Passing H, Bender R, et al. A general regression procedure for method transformation. Application of linear regression procedures for method comparison studies in clinical chemistry, Part III. *J Clin Chem Clin Biochem* 1988 Nov;26(11):783-790.



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim
www.roche.com
+800 5505 6606



For further information, please refer to the appropriate operator's manual for the analyzer concerned, the respective application sheets and the Method Sheets of all necessary components (if available in your country).

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.









CA 19-9 is a registered trademark of Fujirebio Diagnostics, Inc.

Any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the Member State in which the user and/or the patient is established.

The Summary of Safety & Performance Report can be found here:
<https://ec.europa.eu/tools/eudamed>

Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard (for USA: see dialog.roche.com for definition of symbols used):

	Contents of kit
	Analyzers/Instruments on which reagents can be used
	Reagent
	Calibrator
	Volume for reconstitution
	Global Trade Item Number

COBAS, COBAS E, ELECSYS and PRECICONTROL are trademarks of Roche.

All other product names and trademarks are the property of their respective owners.

Additions, deletions or changes are indicated by a change bar in the margin.

© 2021, Roche Diagnostics