

Elecsys CMV IgG

cobas®

REF			SYSTEM
09118551190	09118551500	300	cobas e 402 cobas e 801

English

System information

Short name	ACN (application code number)
CMVIGG	10218

Please note

The measured CMV IgG value of a patient's sample can vary depending on the testing procedure used. The laboratory finding must therefore always contain a statement on the CMV IgG assay method used. CMV IgG values determined on patient samples by different testing procedures cannot be directly compared with one another and could be the cause of erroneous medical interpretations. Therefore, the results reported by the laboratory to the physician should include: "The following results were obtained with the Elecsys CMV IgG assay. Results from assays of other manufacturers cannot be used interchangeably."

Intended use

Immunoassay for the in vitro quantitative determination of IgG antibodies to cytomegalovirus in human serum and plasma.

Results with this assay are used to indicate past or recent infection with CMV.

The electrochemiluminescence immunoassay "ECLIA" is intended for use on **cobas e** immunoassay analyzers.

Summary

The Elecsys CMV IgG assay is intended to be used as an aid to diagnose past or recent infection with CMV in suspected patients and pregnant women. In addition, the assay is intended to be used for screening including the use as the first-line assay of individual human donors of blood, blood components, cells, tissue, and organs, when donor samples are obtained while the donor's heart is still beating and of cadaveric blood specimens (specimens collected post-mortem, non-heart-beating). The use of cadaveric blood specimens has been established according to Paul-Ehrlich-Institut (PEI) recommendation.¹

Cytomegalovirus (CMV), a member of the herpes virus family, is ubiquitous in all human populations, causing infections which are followed by life-long latency in the host with occasional reactivations.^{2,3} The seroprevalence of antibodies in adults ranges from 40-100 % with inverse correlation to socioeconomic status.^{2,3,4,5} CMV is transmitted through body fluids, including blood, genital secretions and breast milk. Saliva and urine of infected individuals also represent a prominent source of infection, and children, especially those attending day care facilities, are a major vector for viral spread.^{3,4,6,7,8} In immunocompetent individuals primary CMV infection is usually mild or asymptomatic.^{3,7} Patients commonly present with a mononucleosis-like syndrome, including fever, sore throat, cervical lymphadenopathy, malaise, headache, muscle ache and joint pains.^{3,4,6,7,9} During pregnancy, CMV can cause congenital infection which may result in permanent physical and/or neurological sequelae in the child.⁷ CMV infection can be primary, i.e. newly acquired, or secondary, i.e. due to reactivation of the latent virus or re-infection with a different viral strain.^{4,7} Primary CMV infection is reported in 1-4 % of seronegative women during pregnancy and the risk of transmission to the fetus is estimated to be about 30-40 %.^{4,6} Reactivation of CMV infection during pregnancy is reported in 10-30 % of seropositive women and, in this circumstance, the risk of transmission of the virus is about 1-3 %.^{4,6,7} Overall, prenatal CMV infection occurs in 0.6-0.7 % of all life births in the developed world.^{6,7,10,11} The majority of babies born with congenital CMV infection are asymptomatic at birth.^{10,12,13} Of these 5-15 % still develop irreversible impairments, most frequently hearing loss, that can occur several months or even years after birth.^{7,10,12,13} For babies symptomatic at birth prognosis is very poor as they are likely to develop severe mental impairment and/or hearing loss.^{7,10,12,13} Different studies have shown that the risk of symptomatic congenital disease in the fetus or newborn infant is high, when maternal primary infection takes place in early pregnancy before week 20 of gestation, and

lower thereafter.^{6,7} The congenital CMV infection caused by recurrent maternal infection seldom leads to symptomatic disease at birth.^{6,7}

At risk for CMV infection and disease are also immunocompromised patients such as transplant recipients and HIV infected patients where the virus can cause life-threatening diseases.^{14,15} The CMV status of transplant donors and recipients is very important, as it will determine prophylactic and pre-emptive treatment strategies against CMV.¹⁶ CMV-negative transplant recipients should receive donations from CMV-negative individuals or leukocyte depleted blood products. During latency, CMV resides in infected cells and the free viral DNA load is usually low. The CMV status can still be determined by testing for CMV IgG antibodies.¹⁷

Within the appropriate clinical context, the first step in diagnosing acute primary CMV infection is most commonly made by the detection of anti-CMV-specific IgG and IgM antibodies.⁷ Samples being reactive for IgM antibodies indicate an acute, recent or reactivated infection.^{3,6,7,15} For further analysis of a primary CMV infection the determination of the CMV IgG avidity is used as an aid.^{3,6,7,15} A positive IgM result in combination with a low avidity index for IgG is a strong indication of a recent primary CMV infection.^{6,7,15} Seroconversion to CMV IgM and IgG may also indicate a recent CMV infection.^{3,4,6,7,15}

Test principle

Sandwich principle. Total duration of assay: 18 minutes.

- 1st incubation: 12 µL of sample, biotinylated recombinant CMV-specific antigens, and CMV-specific recombinant antigens labeled with a ruthenium complex^{a)} form a sandwich complex.
- 2nd incubation: After addition of streptavidin-coated microparticles, the complex becomes bound to the solid phase via interaction of biotin and streptavidin.
- The reaction mixture is aspirated into the measuring cell where the microparticles are magnetically captured onto the surface of the electrode. Unbound substances are then removed with ProCell II M. Application of a voltage to the electrode then induces chemiluminescent emission which is measured by a photomultiplier.
- Results are determined via a calibration curve which is instrument-specifically generated by 2-point calibration and a master curve provided via the **cobas** link.

a) Tris(2,2'-bipyridyl)ruthenium(II)-complex (Ru(bpy)₃²⁺)

Reagents - working solutions

The **cobas e** pack (M, R1, R2) is labeled as CMVIGG.

- M Streptavidin-coated microparticles, 1 bottle, 14.1 mL:
Streptavidin-coated microparticles 0.72 mg/mL; preservative.
- R1 CMV-Ag-biotin, 1 bottle, 18.8 mL:
Biotinylated CMV-specific antigen (recombinant, *E. coli*), > 400 µg/L, 2-(N-morpholino)ethanesulfonic acid (MES) buffer 50 mmol/L, pH 6.5; preservative.
- R2 CMV-Ag-Ru(bpy)₃²⁺, 1 bottle, 18.8 mL:
CMV-specific antigen (recombinant, *E. coli*) labeled with ruthenium complex > 400 µg/L; MES buffer 50 mmol/L, pH 6.5; preservative.

- CMVIGG Cal1 Negative calibrator 1, 1 bottle of 1.0 mL:
Human serum, non-reactive for anti-CMV IgG; buffer; preservative.
- CMVIGG Cal2 Positive calibrator 2, 1 bottle of 1.0 mL:
Human serum, reactive for anti-CMV IgG, approximately 40 U/mL; buffer; preservative.

Precautions and warnings

For in vitro diagnostic use for laboratory professionals. Exercise the normal precautions required for handling all laboratory reagents.

Elecsys CMV IgG

Infectious or microbial waste:

Warning: handle waste as potentially biohazardous material. Dispose of waste according to accepted laboratory instructions and procedures.

Environmental hazards:

Apply all relevant local disposal regulations to determine the safe disposal.

Safety data sheet available for professional user on request.

This kit contains components classified as follows in accordance with the Regulation (EC) No. 1272/2008:



Warning

H317 May cause an allergic skin reaction.

Prevention:

P261 Avoid breathing mist or vapours.

P272 Contaminated work clothing should not be allowed out of the workplace.

P280 Wear protective gloves.

Response:

P333 + P313 If skin irritation or rash occurs: Get medical advice/attention.

P362 + P364 Take off contaminated clothing and wash it before reuse.

Disposal:

P501 Dispose of contents/container to an approved waste disposal plant.

Hazardous components:

- 2-methyl-2H-isothiazol-3-one hydrochloride

Product safety labeling follows EU GHS guidance.

Contact phone: all countries: +49-621-7590

All human material should be considered potentially infectious. All products derived from human blood are prepared exclusively from the blood of donors tested individually and shown to be free from HBsAg and antibodies to HCV and HIV. The testing methods use assays that have been approved or cleared by the FDA or that are in compliance with the legal rules of the European Union (IVDR 2017/746/EU, IVDD 98/79/EC, Annex II, List A). However, as no testing method can rule out the potential risk of infection with absolute certainty, the material should be handled with the same level of care as a patient specimen. In the event of exposure, the directives of the responsible health authorities should be followed.^{18,19}

The serum containing anti-CMV IgG (CMVIGG Cal2) was 0.2 micron filtrated.

Avoid foam formation in all reagents and sample types (specimens, calibrators and controls).

Reagent handling

The reagents (M, R1, R2) in the kit are ready-for-use and are supplied in **cobas e** packs.

Calibrators:

The calibrators are supplied ready-for-use in bottles compatible with the system.

Unless the entire volume is necessary for calibration on the analyzer, transfer aliquots of the ready-for-use calibrators into empty snap-cap bottles (CalSet Vials). Attach the supplied labels to these additional bottles. Store the aliquots at 2-8 °C for later use.

Perform **only one** calibration procedure per aliquot.

All information required for correct operation is available via the **cobas** link.

Storage and stability

Store at 2-8 °C.

Do not freeze.

Store the **cobas e** pack **upright** in order to ensure complete availability of the microparticles during automatic mixing prior to use.

Stability of the cobas e pack:	
unopened at 2-8 °C	up to the stated expiration date
on the analyzers	16 weeks

Stability of the calibrators:	
unopened at 2-8 °C	up to the stated expiration date
after opening at 2-8 °C	16 weeks
on the analyzers at 20-25 °C	use only once

Store calibrators **upright** in order to prevent the calibrator solution from adhering to the snap-cap.

Specimen collection and preparation

Specimen collected from living patients, blood donors, or individual organ, tissue or cell donors may be used, including donor samples obtained while the donor's heart is still beating.

Performance for the use of cadaveric blood specimens (specimens collected post-mortem, non-heart-beating) was established according to Paul-Ehrlich-Institut recommendation¹ with samples obtained within 24 hours after death.²⁰ Qualitative differences of neat (non-reactive) or spiked (reactive) specimens from cadaveric compared to living donors were not observed.

Criterion: Mean value of cadaveric specimens compared to specimens from living donors within a recovery of 75-125 %.

Only the specimens listed below were tested and found acceptable.

Serum collected using standard sampling tubes or tubes containing separating gel.

Li-heparin, Na-heparin, K₂-EDTA, K₃-EDTA, Na-Citrate, CPD, CP2D, CPDA-1 and ACDA plasma.

Criterion: Mean recovery of serum value: negative/borderline samples $\pm$ 0.2 U/mL; reactive samples: 80-120 % for samples containing solid anticoagulants and 55-95 % for samples containing liquid anticoagulants.

Sampling devices containing liquid anticoagulants have a dilution effect resulting in lower values (U/mL) for individual patient specimens.

In order to minimize dilution effects it is essential that respective sampling devices are filled completely according to manufacturer's instructions.

Stability:

For living patients and donor specimens obtained while the donor's heart is still beating: Stable for 7 days at 20-25 °C, 28 days at 2-8 °C, 6 months at -20 °C ($\pm$ 5 °C). The samples may be frozen 5 times.

For cadaveric specimens: Stable for 2 days at 20-25 °C, 7 days at 2-8 °C. The samples may be frozen 3 times.

The sample types listed were tested with a selection of sample collection tubes or systems that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube/collection system manufacturer.

Specimens should not be subsequently altered with additives (e.g. biocides, anti-oxidants or substances that could possibly change the pH or ionic strength of the sample) in order to avoid erroneous findings.

Pooled samples and other artificial material may have different effects on different assays and thus may lead to discrepant findings.

Centrifuge samples containing precipitates and thawed samples before performing the assay.

Do not use post-mortem samples collected later than 24 hours after last heart beat.

If derived from living donors, heat-inactivated and reconstituted lyophilized samples can be used.

Ensure the samples and calibrators are at 20-25 °C prior to measurement.

Due to possible evaporation effects, samples and calibrators on the analyzers should be analyzed/measured within 2 hours.

Elecsys CMV IgG

The performance of the Elecsys CMV IgG assay has not been established with body fluids other than serum and plasma.

Materials provided

See "Reagents – working solutions" section for reagents.

- 2 x 6 bottle labels

Materials required (but not provided)

- [REF] 04784600190, PreciControl CMV IgG, 16 x 1.0 mL
- [REF] 11776576322, CalSet Vials, 2 x 56 empty snap-cap bottles
- [REF] 07299001190, Diluent Universal, 36 mL sample diluent
- General laboratory equipment

- cobas e** analyzer

Additional materials for **cobas e** 402 and **cobas e** 801 analyzers:

- [REF] 06908799190, ProCell II M, 2 x 2 L system solution
- [REF] 04880293190, CleanCell M, 2 x 2 L measuring cell cleaning solution
- [REF] 07485409001, Reservoir Cup, 8 cups to supply ProCell II M and CleanCell M
- [REF] 06908853190, PreClean II M, 2 x 2 L wash solution
- [REF] 05694302001, Assay Tip/Assay Cup tray, 6 magazines x 6 magazine stacks x 105 assay tips and 105 assay cups, 3 wasteliners
- [REF] 07485425001, Liquid Flow Cleaning Cup, 2 adaptor cups to supply ISE Cleaning Solution/Elecsys SysClean for Liquid Flow Cleaning Detection Unit
- [REF] 07485433001, PreWash Liquid Flow Cleaning Cup, 1 adaptor cup to supply ISE Cleaning Solution/Elecsys SysClean for Liquid Flow Cleaning PreWash Unit
- [REF] 11298500316, ISE Cleaning Solution/Elecsys SysClean, 5 x 100 mL system cleaning solution

Assay

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

Resuspension of the microparticles takes place automatically prior to use.

Place the cooled (stored at 2-8 °C) **cobas e** pack on the reagent manager. Avoid foam formation. The system automatically regulates the temperature of the reagents and the opening/closing of the **cobas e** pack.

Calibrators:

Place the calibrators in the sample zone.

Read in all the information necessary for calibrating the assay.

Calibration

Traceability: This method has been standardized against the internal Roche standard for CMV IgG.

The predefined master curve is adapted to the analyzer using CMVIGG Cal1 and CMVIGG Cal2.

Calibration frequency: Calibration must be performed once per reagent lot using CMVIGG Cal1, CMVIGG Cal2 and fresh reagent (i.e. not more than 24 hours since the **cobas e** pack was registered on the analyzer).

Calibration interval may be extended based on acceptable verification of calibration by the laboratory.

Renewed calibration is recommended as follows:

- every 12 weeks when using the same reagent lot
- every 28 days when using the same **cobas e** pack on the analyzer
- as required: e.g. quality control findings outside the defined limits

Quality control

Use Elecsys PreciControl CMV IgG or other suitable controls for routine quality control procedures.

Controls for the various concentration ranges should be run individually at least once every 24 hours when the test is in use, once per **cobas e** pack, and following each calibration.

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined

limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

If necessary, repeat the measurement of the samples concerned.

Follow the applicable government regulations and local guidelines for quality control.

Calculation

The analyzer automatically calculates the analyte concentration of each sample in U/mL, and adds the qualitative interpretation.

Interpretation of the results

Numeric result	Result message	Interpretation/further steps
< 0.5 U/mL	Non-reactive	Not infected with CMV and therefore susceptible to primary infection.
≥ 0.5 to < 1.0 U/mL	Borderline	Samples should be retested. In case the result is still borderline, a second sample should be collected (e.g. within 2 weeks) and testing should be repeated.
≥ 1.0 U/mL	Reactive	Positive for CMV IgG-specific antibodies indicating either acute or past infection. Such individuals are potentially at risk of transmitting the virus (e.g. mother to fetus) but are at current not necessarily contagious.*

* For the diagnosis of acute CMV infection further tests have to be performed e.g. CMV IgM and CMV IgG avidity. A positive IgM result in combination with a low avidity index for IgG is a strong indication of a primary CMV infection within the last 4 months.

The diagnosis may be supported by a significant increase of the CMV IgG antibody titer from a first to a second sample taken e.g. within 3-4 weeks. Note: A borderline or low positive result may already indicate an early acute CMV infection (also if the sample is non-reactive for CMV IgM antibodies).

The anti-CMV IgG results in a given specimen, as determined by assays from different manufacturers, can vary due to differences in assay and reagent methods. Therefore, the results reported by the laboratory to the physician should include: "The following results were obtained with the Elecsys CMV IgG assay. Results from assays of other manufacturers cannot be used interchangeably."

Limitations - interference

The effect of the following endogenous substances and pharmaceutical compounds on assay performance was tested. Interferences were tested up to the listed concentrations and no impact on results was observed.

Endogenous substances

Compound	Concentration tested
Bilirubin	≤ 1129 µmol/L or ≤ 66 mg/dL
Hemoglobin	≤ 0.310 mmol/L or ≤ 500 mg/dL
Intralipid	≤ 2000 mg/dL
Biotin	≤ 4912 nmol/L or ≤ 1200 ng/mL
Rheumatoid factors	≤ 1500 IU/mL

Criterion: Recovery of positive samples within ± 20 % of serum value.

A negative test result does not completely rule out the possibility of an infection with CMV. Individuals may not exhibit any detectable IgG antibodies at the early stage of acute infection.

The detection of CMV-specific IgG antibodies in a single sample indicates a previous exposure to CMV but is not always sufficient to distinguish between an acute or latent infection (irrespective of the level of the IgG antibody titer).

Elecsys CMV IgG



In rare cases of primary CMV infection IgG antibody may be present before a specific IgM antibody response is detected. It is recommended that a follow-up sample be tested after 2 weeks. If the CMV IgG antibody titer remains stable, a primary infection can be excluded.^{21,22}

Elecsys CMV IgG results should be used in conjunction with the patient's medical history, clinical symptoms and other laboratory tests, e.g. CMV-specific IgM results, CMV IgG avidity results.

The results in HIV patients, in patients undergoing immunosuppressive therapy, or in patients with other disorders leading to immune suppression, should be interpreted with caution.

Specimens from neonates, cord blood, pretransplant patients or body fluids other than serum and plasma, such as urine, saliva or amniotic fluid have not been tested.

There is no high-dose hook effect at CMV IgG concentrations up to 2500 U/mL.

Pharmaceutical substances

In vitro tests were performed on 17 commonly used pharmaceuticals. No interference with the assay was found.

In addition, the following special anti-viral drugs were tested. No interference with the assay was found.

Special anti-viral drugs

Drug	Concentration tested
Ganciclovir	≤ 800 mg/L
Valganciclovir	≤ 900 mg/L

In rare cases, interference due to extremely high titers of antibodies to immunological components, streptavidin or ruthenium can occur. These effects are minimized by suitable test design.

For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.

Limits and ranges

Measuring range

0.25-500 U/mL (defined by the Limit of Detection and the maximum of the master curve). Values below the Limit of Blank are reported as < 0.15 U/mL. Values above the Limit of Blank but below the Limit of Detection will not be flagged by the instrument. Values above the measuring range are reported as > 500 U/mL (or up to 10000 U/mL for 20-fold diluted samples).

Lower limits of measurement

Limit of Blank, Limit of Detection

Limit of Blank = 0.15 U/mL

Limit of Detection = 0.25 U/mL

The Limit of Blank and Limit of Detection were determined in accordance with the CLSI (Clinical and Laboratory Standards Institute) EP17-A2 requirements.

The Limit of Blank is the 95th percentile value from $n \geq 60$ measurements of analyte-free samples over several independent series. The Limit of Blank corresponds to the concentration below which analyte-free samples are found with a probability of 95 %.

The Limit of Detection is determined based on the Limit of Blank and the standard deviation of low concentration samples. The Limit of Detection corresponds to the lowest analyte concentration which can be detected (value above the Limit of Blank with a probability of 95 %).

Dilution

Samples with anti-CMV IgG concentrations above the measuring range can be diluted with Diluent Universal. The recommended dilution is 1:20 (either automatically by the analyzers or manually). The concentration of the diluted sample must be ≥ 15 U/mL.

After manual dilution, multiply the result by the dilution factor.

After dilution by the analyzers, the software automatically takes the dilution into account when calculating the sample concentration.

Manual dilution can also be made with human serum negative for CMV IgG.

Note: Antibodies to CMV are heterogeneous. This may lead to non-linear dilution behavior.

Expected values

The prevalence of IgG antibodies to CMV varies considerably depending upon geographical location and socioeconomic status of the population studied.

The Elecsys CMV IgG assay was used to test 616 samples from clinical routine in Germany (site 1) and 520 samples from clinical routine in Israel (site 2). Out of these 334 (54.2 %, Germany) and 415 samples (79.8 %, Israel) were found positive or indeterminate with the Elecsys CMV IgG assay.

A distribution of these values is given in the following table:

	Site 1, Germany, n = 616		Site 2, Israel, n = 520	
U/mL	N	% of total	N	% of total
< 0.5	282	45.8	105	20.2
0.5-< 1	4	0.6	2	0.4
≥ 1-< 10	15	2.4	2	0.4
10-< 100	62	10.1	71	13.7
100-< 300	91	14.8	114	21.9
300-< 500	65	10.6	84	16.2
≥ 500	97	15.7	142	27.3

Each laboratory should investigate the transferability of the expected values to its own patient population and if necessary determine its own reference ranges.

Specific performance data

Representative performance data on the analyzers are given below. Results obtained in individual laboratories may differ.

Precision

Precision was determined using Elecsys reagents, samples and controls in a protocol (EP05-A3) of the CLSI (Clinical and Laboratory Standards Institute): 2 runs per day in duplicate each for 21 days ($n = 84$). The following results were obtained:

cobas e 402 and cobas e 801 analyzers					
		Repeatability		Intermediate precision	
Sample	Mean U/mL	SD U/mL	CV %	SD U/mL	CV %
HSP ^{b)} , low positive	1.08	0.022	2.0	0.026	2.4
HSP, medium positive	267	3.61	1.4	4.67	1.8
HSP, high positive	405	3.87	1.0	7.66	1.9
PC ^{c)} CMV IgG 1	1.32	0.013	1.0	0.023	1.7
PC CMV IgG 2	24.1	0.207	0.9	0.412	1.7

b) HSP = human specimen (serum/plasma)

c) PC = PreciControl

Method comparison

a) A comparison of the Elecsys CMV IgG assay, [REF] 09118551190 (cobas e 801 analyzer; y), with the Elecsys CMV IgG assay, [REF] 07027117190 (cobas e 801 analyzer; x), gave the following correlations (U/mL):

Number of samples measured: 274

Passing/Bablok²³ Linear regression
 $y = 0.892x - 0.0000000$ $y = 0.863x + 0.421$
 $r = 0.966$ $r = 0.995$

The sample concentrations were between 0 and 451 U/mL.

b) A comparison of the Elecsys CMV IgG assay, [REF] 09118551190 (cobas e 402 analyzer; y), with the Elecsys CMV IgG assay, [REF] 09118551190 (cobas e 801 analyzer; x), gave the following correlations (U/mL):

Number of samples measured: 207

Elecsys CMV IgG

Passing/Bablok²³

$$y = 0.929x + 0.000000$$

$$r = 0.986$$

Linear regression

$$y = 0.942x - 0.682$$

$$r = 0.996$$

The sample concentrations were between 0 and 391 U/mL.

A total of 1668 fresh samples obtained from clinical routine (blood donors and pregnancy testing) were tested with the Elecsys CMV IgG assay at 3 different sites in comparison to commercially available CMV IgG assays.

Agreement to comparative assays

Site	N	Agreement ^{d)} %	Concordant reactive	Concordant non-reactive	Discrepant
1 ^{e)}	532	98.9	206	320	6
2 ^{f)}	616	96.8	317	279	20
3 ^{g)}	520	99.4	414	103	3

d) All indeterminate samples counted positive.

e) 4 samples were found indeterminate with the Elecsys CMV IgG assay and negative with the comparison assay; 2 samples were found discrepant negative with the Elecsys CMV IgG assay.

f) 13 samples were found discrepant positive with the Elecsys CMV IgG assay; 4 samples were found indeterminate with the Elecsys CMV IgG assay and negative with the comparison assay; 2 samples were found discrepant negative with the Elecsys CMV IgG assay; 1 sample was found negative with the Elecsys CMV IgG assay and indeterminate with the comparison assay.

g) 1 sample was found indeterminate with the Elecsys CMV IgG assay and negative with the comparison assay; 2 samples were found discrepant negative with the Elecsys CMV IgG assay.

Analytical specificity

439 potentially cross reacting samples were tested with the Elecsys CMV IgG assay and a comparison CMV IgG assay comprising the following specimens:

- containing antibodies against HBV**, HAV, HCV*, HIV, HTLV, EBV**, HSV*, VZV**, Parvo B19***, Rubella, Treponema pallidum**, Toxoplasma gondii**
- containing autoantibodies*** (ANA, anti-tissue, RF)

An overall agreement of 96.6 % (422/437) was found in these specimens with the Elecsys CMV IgG assay and the comparison test. 110 samples were found concordantly negative and 312 samples were found positive. 2 samples were found indeterminate either with the Elecsys CMV IgG assay or the comparison test.

* HSV, HCV: 2 discordant samples were found in each group.

** HBV, EBV, VZV, Treponema pallidum, Toxoplasma gondii: 1 discordant sample was found in each group.

*** Parvo B19, autoantibodies: 3 discordant samples were found in each group.

Agreement in primary infections

A total of 368 frozen samples from pregnant women (acute, recent and late phase) including sequential and single samples analyzed by commercially available CMV IgG assays were tested with the Elecsys CMV IgG assay at 4 different sites.

Agreement to comparative assays

Site	N	Agreement ^{h)} %	Concordant reactive	Concordant non-reactive	Discrepant
1 ⁱ⁾	181	96.1	172	2	7
2 ^{j)}	57	96.5	52	3	2
3	40 ^{k)}	97.5	39	0	1
	36 ^{l)}	94.4	34	0	2
4 ^{m)}	54	90.7	43	6	5

h) All indeterminate samples counted positive.

i) 6 samples were found discordant positive with the Elecsys CMV IgG assay; 1 sample was found discordant negative with the Elecsys CMV IgG assay.

j) 1 sample was found discordant positive with the Elecsys CMV IgG assay; 1 sample was found discordant negative with the Elecsys CMV IgG assay.

k) 1 sample was found discordant negative with the Elecsys CMV IgG assay.

l) 1 sample was found discordant negative with the Elecsys CMV IgG assay and indeterminate with the comparison assay.

m) 4 samples were found discordant negative with the Elecsys CMV IgG assay; 1 sample was found discordant indeterminate with the Elecsys CMV IgG assay; the comparison assay was negative.

Agreement in past infection

A total of 158 frozen samples from pregnant women with past CMV infection analyzed by commercially available CMV IgG assays were tested with the Elecsys CMV IgG assay at 4 different sites. A 100 % agreement for the Elecsys CMV IgG assay was found with the competitor assays.

Agreement in pre-selected negative samples

A total of 161 frozen samples from pregnant women in which a CMV infection was excluded and analyzed by commercially available CMV IgG assays were tested with the Elecsys CMV IgG assay at 4 different sites. In 3 sites a 100 % agreement of the Elecsys CMV IgG assay was found with the competitor assays, whereas at site 4 the Elecsys CMV IgG assay showed 1 discrepant positive and 1 discrepant indeterminate result (agreement 96 %).

References

- Proposal for the Validation of Anti-HIV-1/2 or HIV Ag/Ab Combination Assays, anti-HCV-Assays, HBsAg and Anti-HBc assays for Use with Cadaveric Samples; PEI 08/05/2014.
- Cannon MJ, Schmid DS, Hyde TB. Review of cytomegalovirus seroprevalence and demographic characteristics associated with infection. Rev Med Virol 2010;20:202-213.
- Johnson J, Anderson B, Pass RF. Prevention of maternal and congenital cytomegalovirus infection. Clin Obstet Gynecol 2012;55:521-530.
- Yinon Y, Farine D, Yudin MH. Screening, diagnosis, and management of cytomegalovirus infection in pregnancy. Obstet Gynecol Surv 2010;65:736-743.
- Fowler K, Mucha J, Neumann M et al. A systematic literature review of the global seroprevalence of cytomegalovirus: possible implications for treatment, screening, and vaccine development. BMC Public Health 2022; 22: 1659
- Buonsenso D, Serranti D, Gargiullo L, et al. Congenital cytomegalovirus infection: current strategies and future perspectives. Eur Rev Med Pharmacol Sci 2012;16:919-935.
- Coll O, Benoist G, Ville Y, et al. WAPM Perinatal Infections Working Group. Guidelines on CMV congenital infection. J Perinat Med 2009;37:433-445.
- Joseph S, Beliveau C, Muecke C, et al. Cytomegalovirus as an occupational risk in daycare educators. Paediatr Child Health 2006;11:401-407.
- Lazzarotto T, Guerra B, Lanari M, et al. New advances in the diagnosis of congenital cytomegalovirus infection. J Clin Virol 2008;41:192-197.
- Lombardi G, Garofoli F, Stronati M. Congenital cytomegalovirus infection: treatment, sequelae and follow-up. J Matern Fetal Neonatal Med 2010;23Suppl3:45-48.
- Ssentongo P, Hehrly C, Birungi P, et al. Congenital Cytomegalovirus Infection Burden and Epidemiologic Risk Factors in Countries With Universal Screening: A Systematic Review and Meta-analysis. JAMA Netw Open. 2021;4(8):e2120736.
- Elliott SP. Congenital cytomegalovirus infection: an overview. Infect Disord Drug Targets 2011;11:432-436.
- Lazzarotto T, Guerra B, Gabrielli L, et al. Update on the prevention, diagnosis and management of cytomegalovirus infection during pregnancy. Clin Microbiol Infect 2011;17:1285-1293.
- Plosa EJ, Esbenshade JC, Fuller MP, et al. Cytomegalovirus infection. Pediatr Rev 2012;33:156-163.
- Ross SA, Novak Z, Pati S, et al. Overview of the diagnosis of cytomegalovirus infection. Infect Disord Drug Targets 2011;11:466-474.
- EDQM, Council of Europe. Guide to the Quality and Safety of Organs for Transplantation. 8th ed. Strasbourg: EDQM. [Internet; updated 2022; cited 2023 Sept 8]. Available from: <https://freepub.edqm.eu/publications/PUBSD-88/detail>
- Rawlinson WD, Boppana SB, Fowler KB, et al. Congenital cytomegalovirus infection in pregnancy and the neonate: consensus recommendations for prevention, diagnosis, and therapy. Lancet Infect Dis. 2017 Jun;17(6):e177-e188.

Elecsys CMV IgG



- 18 Occupational Safety and Health Standards: Bloodborne pathogens. (29 CFR Part 1910.1030). Fed. Register.
- 19 Directive 2000/54/EC of the European Parliament and Council of 18 September 2000 on the protection of workers from risks related to exposure to biological agents at work.
- 20 Commission Directive 2006/17/EC of 8 February 2006 implementing Directive 2004/23/EC of the European Parliament and of the Council as regards certain technical requirements for the donation, procurement and testing of human tissues and cells.
- 21 Revello MG, Germa G. Diagnosis and Management of Human Cytomegalovirus Infection in the Mother, Fetus, and Newborn Infant. Clin Microbiol Rev 2002;15(4):680-715.
- 22 Weber B, Fall EM, Berger A, et al. Screening of blood donors for human cytomegalovirus (HCMV) IgG antibody with an enzyme immunoassay using recombinant antigens. J Clin Virol 1999;14:173-181.
- 23 Bablok W, Passing H, Bender R, et al. A general regression procedure for method transformation. Application of linear regression procedures for method comparison studies in clinical chemistry, Part III. J Clin Chem Clin Biochem 1988 Nov;26(11):783-790.

For further information, please refer to the appropriate user guide or operator's manual for the analyzer concerned, the respective application sheets and the Method Sheets of all necessary components (if available in your country).

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the Member State in which the user and/or the patient is established.

The Summary of Safety & Performance Report can be found here: <https://ec.europa.eu/tools/eudamed>

Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard (for USA: see navifyportal.roche.com for definition of symbols used):

	Contents of kit
	Analyzers/Instruments on which reagents can be used
	Reagent
	Calibrator
	Volume for reconstitution
	Global Trade Item Number

Rx only For USA: Caution: Federal law restricts this device to sale by or on the order of a physician.

COBAS, NAVIFY, ELECSYS and PRECICONTROL are trademarks of Roche. INTRALIPID is a trademark of Fresenius Kabi AB.

All other product names and trademarks are the property of their respective owners.

Additions, deletions or changes are indicated by a change bar in the margin.

© 2024, Roche Diagnostics



Roche Diagnostics GmbH
Sandhofer Strasse 116
68305 Mannheim, Germany
www.roche.com

+800 5505 6606

