

REF	CONTENT	Analyzer(s) on which cobas c pack(s) can be used
03289923190	Bicarbonate Liquid (250 tests)	System-ID 07 6725 5 cobas c 311, cobas c 501/502

Materials required (but not provided):

20751995190	Ammonia/Ethanol/CO2 Calibrator (2 x 4 mL)	Code 688	
20752401190	Ammonia/Ethanol/CO2 Control Normal (5 x 4 mL)	Code 100	
20753009190	Ammonia/Ethanol/CO2 Control Abnormal (5 x 4 mL)	Code 101	
12149435122	Precinorm U plus (10 x 3 mL)	Code 300	
12149435160	Precinorm U plus (10 x 3 mL, for USA)	Code 300	
12149443122	Precipath U plus (10 x 3 mL)	Code 301	
12149443160	Precipath U plus (10 x 3 mL, for USA)	Code 301	

English

System information

For **cobas c 311/501** analyzers:

CO2-L: ACN 156

SCO2L: ACN 763 (STAT, reaction time: 5) (for **cobas c 501** analyzer)

SCO2L: ACN 763 (STAT, reaction time: 4) (for **cobas c 311** analyzer)

For **cobas c 502** analyzer:

CO2-L: ACN 8156

SCO2L: ACN 8763 (STAT, reaction time: 5)

Intended use

In vitro test for the quantitative determination of bicarbonate (HCO_3^-) in human serum and plasma on **cobas c** systems.

Summary

Bicarbonate is the second largest fraction of the anions in plasma. Included in this fraction are the bicarbonate (HCO_3^-) and carbonate (CO_3^{2-}) ions, as well as the carbamino compounds. At the physiological pH of blood, the concentration of carbonate is 1/1000 that of bicarbonate. The carbamino compounds are also present in such low quantities that they are generally not mentioned specifically.


Several different methods for the determination of bicarbonate in serum and plasma have been reported. Most of these procedures utilize acidification of the sample and conversion of all carbon dioxide forms to CO_2 gas.¹ The amount of gas formed is measured by manometric or volumetric devices, ion selective electrodes, or spectrophotometric techniques.^{2,3} These methods are either cumbersome, time-consuming, technique-oriented, and/or require special equipment.

Enzymatic procedures using phosphoenolpyruvate carboxylase (PEPC) have been described.^{4,5}


The bicarbonate content of serum or plasma is a significant indicator of electrolyte dispersion and anion deficit. Together with pH determination, bicarbonate measurements are used in the diagnosis and treatment of numerous potentially serious disorders associated with acid-base imbalance in the respiratory and metabolic systems.

Test principle

Bicarbonate reacts with phosphoenolpyruvate (PEP) in the presence of PEPC to produce oxaloacetate and phosphate:



The above reaction is coupled with one involving the transfer of a hydrogen ion from NADH analog to oxaloacetate using MDH.



The resultant consumption of NADH analog causes a decrease in absorbance, which is proportional to the concentration of bicarbonate in the sample being assayed.

Reagents - working solutions

R1 Phosphoenolpyruvate: ≥ 40 mmol/L; NADH analog: ≥ 2 mmol/L; MDH (porcine): ≥ 314.3 $\mu\text{kat/L}$; PEPC (microbial): ≥ 30.8 $\mu\text{kat/L}$

R1 is in position B.

Precautions and warnings

For in vitro diagnostic use for health care professionals. Exercise the normal precautions required for handling all laboratory reagents.

Infectious or microbial waste:

Warning: handle waste as potentially biohazardous material. Dispose of waste according to accepted laboratory instructions and procedures.

Environmental hazards:

Apply all relevant local disposal regulations to determine the safe disposal.

Safety data sheet available for professional user on request.

For USA: Caution: Federal law restricts this device to sale by or on the order of a physician.

Reagent handling

Ready for use

Storage and stability

CO2-L

Shelf life at 2-8 °C: See expiration date on **cobas c** pack label.

On-board in use and refrigerated on the analyzer: 6 weeks

Specimen collection and preparation

For specimen collection and preparation only use suitable tubes or collection containers.

Only the specimens listed below were tested and found acceptable.

Serum

Plasma: Li-heparin plasma

The preferred specimen is from venous blood collected anaerobically in the usual manner for bicarbonate analysis. Bicarbonate content in uncapped tubes decreases approximately 4 mmol/L after one hour.⁶ It has been reported that alkalinized serum stored in open cups is stable for up to 4 hours.⁶

Storage of serum at -20 °C or -80 °C for up to 6 months had no significant effect.⁷

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube manufacturer.

Centrifuge samples containing precipitates before performing the assay.

See the limitations and interferences section for details about possible sample interferences.

Sample stability claims were established by experimental data by the manufacturer or based on reference literature and only for the

CO2-L

Bicarbonate Liquid



temperatures/time frames as stated in the method sheet. It is the responsibility of the individual laboratory to use all available references and/or its own studies to determine specific stability criteria for its laboratory.

Stability: 7 days at 4-8 °C⁸
40 hours at 15-25 °C^{9,10}

Separate from erythrocytes and store tightly stoppered.

Materials provided

See "Reagents – working solutions" section for reagents.

Materials required (but not provided)

See "Order information" section

General laboratory equipment

Assay

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

The performance of applications not validated by Roche is not warranted and must be defined by the user.

Application for serum and plasma

cobas c 311 test definition

Assay type	2-Point Rate		
Reaction time / Assay points	10 / 2-18 (STAT 4 / 2-18)		
Wavelength (sub/main)	505/415 nm		
Reaction direction	Decrease		
Unit	mmol/L		
Reagent pipetting	Diluent (H ₂ O)		
R1	50 µL	130 µL	
R2	–	–	
Sample volumes	Sample	Sample dilution	
		Sample	Diluent (H ₂ O)
Normal	2 µL	–	–
Decreased	2 µL	–	–
Increased	2 µL	–	–

cobas c 501 test definition

Assay type	2-Point Rate		
Reaction time / Assay points	10 / 4-29 (STAT 5 / 4-29)		
Wavelength (sub/main)	505/415 nm		
Reaction direction	Decrease		
Unit	mmol/L		
Reagent pipetting	Diluent (H ₂ O)		
R1	50 µL	130 µL	
R2	–	–	
Sample volumes	Sample	Sample dilution	
		Sample	Diluent (H ₂ O)
Normal	2 µL	–	–
Decreased	2 µL	–	–
Increased	2 µL	–	–

cobas c 502 test definition

Assay type	2-Point Rate
Reaction time / Assay points	10 / 4-29 (STAT 5 / 4-29)

Wavelength (sub/main)	505/415 nm		
Reaction direction	Decrease		
Unit	mmol/L		
Reagent pipetting	Diluent (H ₂ O)		
R1	50 µL	130 µL	
R2	–	–	
Sample volumes	Sample	Sample dilution	
		Sample	Diluent (H ₂ O)
Normal	2 µL	–	–
Decreased	2 µL	–	–
Increased	4 µL	–	–

Calibration

Calibrators	S1: H ₂ O S2: Ammonia/Ethanol /CO ₂ Calibrator
Calibration mode	Linear
Calibration frequency	2-point calibration - after reagent lot change - as required following quality control procedures

Calibration interval may be extended based on acceptable verification of calibration by the laboratory.

Traceability: This method has been standardized against a primary standard traceable to NIST.

Quality control

For quality control, use control materials as listed in the "Order information" section. In addition, other suitable control material can be used.

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

Follow the applicable government regulations and local guidelines for quality control.

Calculation

cobas c systems automatically calculate the analyte concentration of each sample.

Limitations – interference

Criterion: Recovery within ± 10 % of initial value at a bicarbonate concentration of 22 mmol/L.

Icterus:¹¹ No significant interference up to an I index of 60 for conjugated and unconjugated bilirubin (approximate conjugated and unconjugated bilirubin concentration: 1026 µmol/L or 60 mg/dL).

Hemolysis:¹¹ No significant interference up to an H index of 600 (approximate hemoglobin concentration: 372.6 µmol/L or 600 mg/dL).

Lipemia (Intralipid):¹¹ No significant interference up to an L index of 1800. There is poor correlation between the L index (corresponds to turbidity) and triglycerides concentration.

Immunoglobulins: No significant interference from immunoglobulins up to a concentration of 35 g/L (233.5 µmol/L) (simulated by human immunoglobulin G).

Drugs: No interference was found at therapeutic concentrations using common drug panels.^{12,13}

In very rare cases, gammopathy, in particular type IgM (Waldenström's macroglobulinemia), may cause unreliable results.¹⁴

For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.

An abnormally elevated concentration of ambient carbon dioxide (CO₂) may occur under certain environmental conditions in the laboratory. The

CO₂-L

Bicarbonate Liquid

cobas®

fluctuating ambient CO₂ concentration may interfere with the CO₂-L assay leading to higher CO₂ results. Under these circumstances, the reduction of the re-calibration interval may become necessary if the laboratory is unable to keep the ambient CO₂ concentration at a normal level by appropriate countermeasures.

ACTION REQUIRED

Special Wash Programming: The use of special wash steps is mandatory when certain test combinations are run together on **cobas c** systems. The latest version of the carry-over evasion list can be found with the NaOH-SMS-SmpCln1+2-SCCS Method Sheets. For further instructions refer to the operator's manual. **cobas c** 502 analyzer: All special wash programming necessary for avoiding carry-over is available via the **cobas** link, manual input is required in certain cases.

Where required, special wash/carry-over evasion programming must be implemented prior to reporting results with this test.

Limits and ranges

Measuring range

2-50 mmol/L

Lower limits of measurement

Lower detection limit of the test:

2 mmol/L

The lower detection limit represents the lowest measurable analyte level that can be distinguished from zero. It is calculated as the value lying 3 standard deviations above that of the lowest standard (standard 1 + 3 SD, repeatability, n = 21).

Expected values¹

22-29 mmol/L

Each laboratory should investigate the transferability of the expected values to its own patient population and if necessary determine its own reference ranges.

Specific performance data

Representative performance data on the analyzers are given below. Results obtained in individual laboratories may differ.

Precision

Precision was determined using human samples and controls in an internal protocol with repeatability (n = 21) and intermediate precision (3 aliquots per run, 1 run per day, 21 days). The following results were obtained:

<i>Repeatability</i>	<i>Mean</i> <i>mmol/L</i>	<i>SD</i> <i>mmol/L</i>	<i>CV</i> <i>%</i>
Ammonia/Ethanol/CO ₂ Control Normal	16.1	0.2	1.0
Ammonia/Ethanol/CO ₂ Control Abnormal	26.5	0.2	0.7
Human serum 1	16.0	0.1	0.8
Human serum 2	27.0	0.2	0.8
<i>Intermediate precision</i>	<i>Mean</i> <i>mmol/L</i>	<i>SD</i> <i>mmol/L</i>	<i>CV</i> <i>%</i>
Ammonia/Ethanol/CO ₂ Control Normal	17.6	0.2	1.3
Ammonia/Ethanol/CO ₂ Control Abnormal	30.5	0.4	1.4
Human serum 3	9.90	0.23	2.3
Human serum 4	26.3	0.3	1.3

The data obtained on **cobas c** 501 analyzer(s) are representative for **cobas c** 311 analyzer(s).

Method comparison

Bicarbonate values for human serum and plasma samples obtained on a **cobas c** 501 analyzer (y) were compared with those determined using the corresponding reagent on a Roche/Hitachi 917 analyzer (x).

Sample size (n) = 73

Passing/Bablok ¹⁵	Linear regression
$y = 1.017x - 0.053 \text{ mmol/L}$	$y = 1.007x + 0.087 \text{ mmol/L}$
$r = 0.976$	$r = 0.998$

The sample concentrations were between 2.54 and 49.9 mmol/L.

The data obtained on **cobas c** 501 analyzer(s) are representative for **cobas c** 311 analyzer(s).

References

- Scott MG, Heusel JW, LeGrys VA, et al. Electrolytes and blood gases, in Tietz NW. Textbook of Clinical Chemistry. 3rd ed. Philadelphia, PA: WB Saunders Co 1999;1065-1066.
- Natelson S. Microtechniques of Clinical Chemistry. Springfield, IL: Charles C Thomas 1975;147.
- Segal MA. A rapid electrotitrimetric method for determining CO₂ combining power in plasma or serum. Am J Clin Pathol 1955;25:1212-1216.
- Wilson W, Jesyk P, Rand R, et al. Use of Vickers discrete analyzer for enzymatic determination of the bicarbonate content of serum. Clin Chem 1973;19(6):640.
- Norris KA, Atkinson AR, Smith WG, et al. Colorimetric enzymatic determination of serum total carbon dioxide, as applied to the Vickers Multichannel 300 discrete analyzer. Clin Chem 1975;21:1093-1101.
- Gambino SR, Schreiber H. The measurement of CO₂ content with the autoanalyzer. A comparison with 3 standard methods and a description of a new method (alkalinization) for preventing loss of CO₂ from open cups. Am J Clin Path 1966;45:406.
- Elfath D, Cooney J, McDaniel R, et al. Effect of frozen storage of serum on the level of 22 chemistry analytes. Clin Chem 1991;37:931.
- Guder WG, da Fonseca-Wollheim F, Heil W, et al. Quality of Diagnostic Samples, in brochure: Recommendations of the Working Group on Preanalytical Quality of the German Society for Clinical Chemistry and Laboratory Medicine, 3rd Ed. 2010.
- Boyanton BL, Blick KE. Stability studies of Twenty-Four Analytes in Human Plasma and Serum. Clin Chem 2002;48/12:2242-2247.
- O'Keane MP, Cunningham SK. Evaluation of three different specimen types (serum, plasma lithium heparin and serum gel separator) for analysis of certain analytes: clinical significance of differences in results and efficiency in use. Clin Chem Lab Med 2006;44(5):662-668.
- Glick MR, Ryder KW, Jackson SA. Graphical Comparisons of Interferences in Clinical Chemistry Instrumentation. Clin Chem 1986;32:470-475.
- Breuer J. Report on the Symposium "Drug effects in Clinical Chemistry Methods". Eur J Clin Chem Clin Biochem 1996;34:385-386.
- Sonntag O, Scholer A. Drug interference in clinical chemistry: recommendation of drugs and their concentrations to be used in drug interference studies. Ann Clin Biochem 2001;38:376-385.
- Bakker AJ, Mücke M. Gammopathy interference in clinical chemistry assays: mechanisms, detection and prevention. Clin Chem Lab Med 2007;45(9):1240-1243.
- Bablok W, Passing H, Bender R, et al. A general regression procedure for method transformation. Application of linear regression procedures for method comparison studies in clinical chemistry, Part III. J Clin Chem Clin Biochem 1988 Nov;26(11):783-790.

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the Member State in which the user and/or the patient is established.

Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard (for USA: see dialog.roche.com for definition of symbols used):

CONTENT	Contents of kit
→	Volume for reconstitution
GTIN	Global Trade Item Number

CO2-L

Bicarbonate Liquid

cobas®

FOR US CUSTOMERS ONLY: LIMITED WARRANTY

Roche Diagnostics warrants that this product will meet the specifications stated in the labeling when used in accordance with such labeling and will be free from defects in material and workmanship until the expiration date printed on the label. THIS LIMITED WARRANTY IS IN LIEU OF ANY OTHER WARRANTY, EXPRESS OR IMPLIED, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR PARTICULAR PURPOSE. IN NO EVENT SHALL ROCHE DIAGNOSTICS BE LIABLE FOR INCIDENTAL, INDIRECT, SPECIAL OR CONSEQUENTIAL DAMAGES.

COBAS, COBAS C, PRECINORM and PRECIPATH are trademarks of Roche.

All other product names and trademarks are the property of their respective owners.

Additions, deletions or changes are indicated by a change bar in the margin.

© 2022, Roche Diagnostics



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim
www.roche.com

+800 5505 6606



Distribution in USA by:
Roche Diagnostics, Indianapolis, IN
US Customer Technical Support 1-800-428-2336