

# Elecsys Estradiol III

REF		$\Sigma$	SYSTEM
06656021190*	06656021500	100	<b>cobas e 411</b>
06656021214*			<b>cobas e 601</b> <b>cobas e 602</b>

\* Some kits shown may not be available in all countries.

## English

### System information

For **cobas e 411** analyzer: test number 1370

For **cobas e 601** and **cobas e 602** analyzers: Application Code Number 223

### Intended use

Immunoassay for the in vitro quantitative determination of estradiol in human serum and plasma.

The electrochemiluminescence immunoassay "ECLIA" is intended for use on Elecsys and **cobas e** immunoassay analyzers.

### Summary

Estrogens are responsible for the development of the secondary female sex characteristics. Together with gestagens they control all the important female reproductive processes.

The biologically most active estrogen is 17 $\beta$ -estradiol. This is a steroid hormone having a molecular weight of 272 daltons.

Estrogens are produced primarily in the ovary (follicle, corpus luteum), but small quantities are also formed in the testes and in the adrenal cortex. During pregnancy, estrogens are mainly formed in the placenta.<sup>1</sup> In human plasma the bulk of estradiol is bound specifically to SHBG (= sex hormone binding globulin) and non-specifically to human serum albumin.<sup>2</sup>

Estrogen secretion is biphasic during the menstrual cycle. The determination of estradiol is utilized clinically in the elucidation of fertility disorders in the hypothalamus-pituitary-gonad axis, gynecomastia, estrogen-producing ovarian and testicular tumors. Further clinical indications are the monitoring of fertility therapy and determining the time of ovulation within the framework of in vitro fertilization (IVF).<sup>1,3</sup>

The Elecsys Estradiol III assay employs a competitive test principle using two monoclonal antibodies specifically directed against 17 $\beta$ -estradiol. Endogenous estradiol released from the sample by mesterolone competes with the added estradiol derivative labeled with a ruthenium complex<sup>a)</sup> for the binding sites on the biotinylated antibody.

a) Tris(2,2'-bipyridyl)ruthenium(II)-complex (Ru(bpy)<sub>3</sub><sup>2+</sup>)

### Test principle

Competition principle. Total duration of assay: 18 minutes.

- 1st incubation: By incubating the sample (25  $\mu$ L) with two estradiol-specific biotinylated antibodies, immunocomplexes are formed, the amount of which is dependent upon the analyte concentration in the sample.
- 2nd incubation: After addition of streptavidin-coated microparticles and an estradiol derivative labeled with a ruthenium complex, the still-vacant sites of the biotinylated antibodies become occupied, with formation of an antibody-hapten complex. The entire complex becomes bound to the solid phase via interaction of biotin and streptavidin.
- The reaction mixture is aspirated into the measuring cell where the microparticles are magnetically captured onto the surface of the electrode. Unbound substances are then removed with ProCell/ProCell M. Application of a voltage to the electrode then induces chemiluminescent emission which is measured by a photomultiplier.
- Results are determined via a calibration curve which is instrument-specifically generated by 2-point calibration and a master curve provided via the reagent barcode or e-barcode.

### Reagents - working solutions

The reagent rackpack is labeled as E2 III.

- M Streptavidin-coated microparticles (transparent cap), 1 bottle, 6.5 mL:  
Streptavidin-coated microparticles 0.72 mg/mL; preservative.

R1 Anti-estradiol-Ab-biotin (gray cap), 1 bottle, 9 mL:

Two biotinylated monoclonal anti-estradiol antibodies (rabbit) 2.5 ng/mL and 4.5 ng/mL; mesterolone 130 ng/mL; MES<sup>b)</sup> buffer 50 mmol/L, pH 6.0; preservative.

R2 Estradiol-peptide~Ru(bpy)<sub>3</sub><sup>2+</sup> (black cap), 1 bottle, 9 mL:

Estradiol derivative, labeled with ruthenium complex 4.5 ng/mL; MES buffer 50 mmol/L, pH 6.0; preservative.

b) MES = 2-morpholino-ethane sulfonic acid

### Precautions and warnings

For in vitro diagnostic use for health care professionals. Exercise the normal precautions required for handling all laboratory reagents.

Infectious or microbial waste:

Warning: handle waste as potentially biohazardous material. Dispose of waste according to accepted laboratory instructions and procedures.

Environmental hazards:

Apply all relevant local disposal regulations to determine the safe disposal.

Safety data sheet available for professional user on request.

For USA: Caution: Federal law restricts this device to sale by or on the order of a physician.

This kit contains components classified as follows in accordance with the Regulation (EC) No. 1272/2008:



### Warning

H317 May cause an allergic skin reaction.

### Prevention:

P261 Avoid breathing dust/fume/gas/mist/vapours/spray.

P272 Contaminated work clothing should not be allowed out of the workplace.

P280 Wear protective gloves.

### Response:

P333 + P313 If skin irritation or rash occurs: Get medical advice/attention.

P362 + P364 Take off contaminated clothing and wash it before reuse.

### Disposal:

P501 Dispose of contents/container to an approved waste disposal plant.

Product safety labeling follows EU GHS guidance.

Contact phone: all countries: +49-621-7590, USA: 1-800-428-2336

Avoid foam formation in all reagents and sample types (specimens, calibrators and controls).

### Reagent handling

The reagents in the kit have been assembled into a ready-for-use unit that cannot be separated.

All information required for correct operation is read in from the respective reagent barcodes.

# Elecsys Estradiol III

## Storage and stability

Store at 2-8 °C.

Do not freeze.

Store the Elecsys reagent kit **upright** in order to ensure complete availability of the microparticles during automatic mixing prior to use.

Stability:	
unopened at 2-8 °C	up to the stated expiration date
after opening at 2-8 °C	12 weeks
on the analyzers	8 weeks

## Specimen collection and preparation

Only the specimens listed below were tested and found acceptable.

Serum collected using standard sampling tubes or tubes containing separating gel.

Li-heparin, K<sub>2</sub>-EDTA and K<sub>3</sub>-EDTA plasma as well as plasma separation tubes.

Criterion: Slope 0.9-1.1 + intercept within  $\pm 10$  pg/mL + coefficient of correlation  $\geq 0.95$ .

For individual samples recovery within 70-130 % of serum value > 100 pg/mL, recovery of  $\pm 20$  pg/mL of serum value  $\leq 100$  pg/mL.

Stable for 24 hours at 20-25 °C, 2 days at 2-8 °C, 6 months at -20 °C ( $\pm 5$  °C). Freeze only once.

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube manufacturer.

Centrifuge samples containing precipitates and frozen samples before performing the assay.

Do not use heat-inactivated samples.

Do not use samples and controls stabilized with azide.

Ensure the samples, calibrators and controls are at 20-25 °C prior to measurement.

Due to possible evaporation effects, samples, calibrators and controls on the analyzers should be analyzed/measured within 2 hours.

Sample stability claims were established by experimental data by the manufacturer or based on reference literature and only for the temperatures/time frames as stated in the method sheet. It is the responsibility of the individual laboratory to use all available references and/or its own studies to determine specific stability criteria for its laboratory.

## Materials provided

See "Reagents – working solutions" section for reagents.

## Materials required (but not provided)

- [REF] 06656048190, Estradiol III CalSet, for 4 x 1.0 mL
- [REF] 11731416190, PreciControl Universal, for 4 x 3.0 mL
- [REF] 11731416160, PreciControl Universal, for 4 x 3.0 mL (for USA)
- [REF] 03609987190, Diluent MultiAssay, 2 x 16 mL sample diluent
- General laboratory equipment
- **cobas e** analyzer

Additional materials for the **cobas e 411** analyzer:

- [REF] 11662988122, ProCell, 6 x 380 mL system buffer
- [REF] 11662970122, CleanCell, 6 x 380 mL measuring cell cleaning solution
- [REF] 11930346122, Elecsys SysWash, 1 x 500 mL washwater additive
- [REF] 11933159001, Adapter for SysClean
- [REF] 11706802001, AssayCup, 60 x 60 reaction cups
- [REF] 11706799001, AssayTip, 30 x 120 pipette tips
- [REF] 11800507001, Clean-Liner

Additional materials for **cobas e 601** and **cobas e 602** analyzers:

- [REF] 04880340190, ProCell M, 2 x 2 L system buffer
- [REF] 04880293190, CleanCell M, 2 x 2 L measuring cell cleaning solution
- [REF] 03023141001, PC/CC-Cups, 12 cups to prewarm ProCell M and CleanCell M before use
- [REF] 03005712190, ProbeWash M, 12 x 70 mL cleaning solution for run finalization and rinsing during reagent change
- [REF] 03004899190, PreClean M, 5 x 600 mL detection cleaning solution
- [REF] 12102137001, AssayTip/AssayCup, 48 magazines x 84 reaction cups or pipette tips, waste bags
- [REF] 03023150001, WasteLiner, waste bags
- [REF] 03027651001, SysClean Adapter M

Additional materials for all analyzers:

- [REF] 11298500316, ISE Cleaning Solution/Elecsys SysClean, 5 x 100 mL system cleaning solution
- [REF] 11298500160, ISE Cleaning Solution/Elecsys SysClean, 5 x 100 mL system cleaning solution (for USA)

## Assay

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

Resuspension of the microparticles takes place automatically prior to use. Read in the test-specific parameters via the reagent barcode. If in exceptional cases the barcode cannot be read, enter the 15-digit sequence of numbers.

**cobas e 601** and **cobas e 602** analyzers: PreClean M solution is necessary.

Bring the cooled reagents to approximately 20 °C and place on the reagent disk (20 °C) of the analyzer. Avoid foam formation. The system automatically regulates the temperature of the reagents and the opening/closing of the bottles.

## Calibration

Traceability: This method has been standardized against CRM 6004a via ID-GC/MS (isotope dilution-gas chromatography/mass spectrometry).<sup>4</sup>

Every Elecsys reagent set has a barcoded label containing specific information for calibration of the particular reagent lot. The predefined master curve is adapted to the analyzer using the relevant CalSet.

**Calibration frequency:** Calibration must be performed once per reagent lot using fresh reagent (i.e. not more than 24 hours since the reagent kit was registered on the analyzer).

Calibration interval may be extended based on acceptable verification of calibration by the laboratory.

Renewed calibration is recommended as follows:

- after 1 month (28 days) when using the same reagent lot
- after 7 days (when using the same reagent kit on the analyzer)
- as required: e.g. quality control findings outside the defined limits

## Quality control

For quality control, use PreciControl Universal.

In addition, other suitable control material can be used.

Controls for the various concentration ranges should be run individually at least once every 24 hours when the test is in use, once per reagent kit, and following each calibration.

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

If necessary, repeat the measurement of the samples concerned.

Follow the applicable government regulations and local guidelines for quality control.

# Elecsys Estradiol III

## Calculation

The analyzer automatically calculates the analyte concentration of each sample (either in pmol/L, pg/mL, ng/L or additionally in nmol/L with **cobas e 601** and **cobas e 602** analyzers).

Conversion factors:                      pmol/L x 0.272 = pg/mL (ng/L)  
     pg/mL x 3.67 = pmol/L  
     pg/mL x 0.00367 = nmol/L

## Limitations - interference

The assay is unaffected by icterus (bilirubin  $\leq$  1129  $\mu$ mol/L or  $\leq$  66 mg/dL), hemolysis (Hb  $\leq$  0.621 mmol/L or  $\leq$  1.0 g/dL), lipemia (Intralipid  $\leq$  1000 mg/dL) and biotin ( $\leq$  147 nmol/L or  $\leq$  36 ng/mL).

Criterion: Recovery within  $\pm$  10 % of initial value.

Samples should not be taken from patients receiving therapy with high biotin doses (i.e. > 5 mg/day) until at least 8 hours following the last biotin administration.

No interference was observed from rheumatoid factors up to a concentration of 1200 IU/mL.

In vitro tests were performed on 17 commonly used pharmaceuticals. No interference with the assay was found.

Erroneous test results may be obtained from samples taken from patients who have been exposed to vaccines containing rabbit serum or when keeping rabbits as pet animals.

Due to the risk of cross reactivity, this assay should not be used when monitoring Estradiol levels in patients being treated with Fulvestrant.

Steroid drugs may interfere with this test.

In rare cases, interference due to extremely high titers of antibodies to analyte-specific antibodies, streptavidin or ruthenium can occur. These effects are minimized by suitable test design.

For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.

## Limits and ranges

### Measuring range

18.4-11010 pmol/L (5-3000 pg/mL) (defined by the Limit of Detection and the maximum of the master curve). Values below the Limit of Detection are reported as < 18.4 pmol/L or < 5 pg/mL. Values above the measuring range are reported as > 11010 pmol/L or > 3000 pg/mL (or up to 110100 pmol/L or 30000 pg/mL for 10-fold diluted samples).

### Lower limits of measurement

*Limit of Blank, Limit of Detection and Limit of Quantitation*

Limit of Blank = 11 pmol/L (3 pg/mL)

Limit of Detection = 18.4 pmol/L (5 pg/mL)

Limit of Quantitation = 91.8 pmol/L (25 pg/mL) with a total allowable error of  $\leq$  30 %

A study was performed for Limit of Quantitation using human serum samples diluted and measured in 6 runs over  $\geq$  3 days on 2 analyzers. At a total allowable error of  $\leq$  30 % Limit of Quantitation was 61.3 pmol/L (16.7 pg/mL).

The Limit of Blank, Limit of Detection and Limit of Quantitation were determined in accordance with the CLSI (Clinical and Laboratory Standards Institute) EP17-A requirements.

The Limit of Blank is the 95<sup>th</sup> percentile value from  $n \geq$  60 measurements of analyte-free samples over several independent series. The Limit of Blank corresponds to the concentration below which analyte-free samples are found with a probability of 95 %.

The Limit of Detection is determined based on the Limit of Blank and the standard deviation of low concentration samples. The Limit of Detection corresponds to the lowest analyte concentration which can be detected (value above the Limit of Blank with a probability of 95 %).

The Limit of Quantitation is defined as the lowest amount of analyte in a sample that can be accurately quantitated with a total allowable error of  $\leq$  30 %.

### Dilution

Samples with estradiol concentrations above the measuring range can be diluted with Diluent MultiAssay. The recommended dilution is 1:10

(automatically by the analyzers). The concentration of the diluted sample must be > 881 pmol/L (> 240 pg/mL).

After dilution by the analyzers, the software automatically takes the dilution into account when calculating the sample concentration.

The endogenous analyte concentration of the diluent (< 220 pmol/L or < 60.0 pg/mL) is not taken into account for dilutions above the measuring range.

### Expected values

The expected ranges were determined by testing specimens drawn from 150 apparently healthy males, 142 apparently healthy, post-menopausal women over the age of 50, and from 412 apparently healthy pregnant women between the ages of 18 and 50 (136 in the first trimester, 140 in the second trimester, and 136 in the third trimester). The expected range for healthy women was determined by collecting blood at multiple time points of one menstrual cycle from 85 apparently healthy subjects with a natural menstrual cycle that were not taking any hormonal contraceptives. A menstrual cycle was defined as the phase between two subsequent menstrual bleedings. Cycle length (29 days) and day of ovulation (day 15) were standardized to account for variation in cycle length within the study population and to enable determination of expected values for further sub-phases. Only ovulatory menstrual cycles were used for value analysis. The following ranges were obtained:

Test subjects	N	2.5th percentile pmol/L (90 % CI*)	Median pmol/L (90 % CI)	97.5th percentile pmol/L (90 % CI)
Healthy men	150	41.4 (22.4-49.0)	90.9 (84.9-97.7)	159 (151-337)
Healthy postmenopausal women				
• Postmenopause	142	< 18.4 (< 18.4-< 18.4)	< 18.4 (< 18.4-19.2)	505 (189-1151)
Healthy pregnant women				
• 1st trimester	136	563 (467-636)	3133 (2703-4004)	11902 (9891-15271)
• 2nd trimester	140	5729 (4173-7457)	28402 (24207-32090)	78098 (69143-92227)
• 3rd trimester	136	31287 (27151-34175)	64684 (62353-68189)	> 110100 (107164-> 110100)

\* CI = confidence interval

Healthy women Cycle Phase	N **	5th percentile pmol/L (90 % CI)	Median pmol/L (90 % CI)	95th percentile pmol/L (90 % CI)
Follicular	85	114 (19.1-135)	198 (188-208)	332 (322-637)
Ovulation	81	222 (98.5-283)	757 (667-944)	1959 (1598-3338)
Luteal	85	222 (159-280)	412 (390-488)	854 (760-1334)

\*\*N = number of patients contributing to the data in this menstrual cycle phase (not number of samples); differences in N per phase are due to cycle standardization procedure

Healthy women Cycle Sub-Phase	N	5th percentile pmol/L (90 % CI)	Median pmol/L (90 % CI)	95th percentile pmol/L (90 % CI)
Early follicular	78	75.5 (< 18.4-78.5)	125 (120-135)	231 (192-283)
Intermediate follicular	83	95.6 (19.1-114)	172 (159-180)	294 (262-695)
Late follicular	84	182 (84-215)	464 (424-519)	858 (711-1337)

# Elecsys Estradiol III

Healthy women Cycle Sub-Phase	N	5th percentile pmol/L (90 % CI)	Median pmol/L (90 % CI)	95th percentile pmol/L (90 % CI)
Ovulation	79	222 (98.5-283)	817 (724-974)	2212 (1598-3338)
Early luteal	85	188 (163-218)	390 (330-412)	658 (608-1394)
Intermediate luteal	81	244 (157-334)	505 (445-568)	1123 (942-1538)
Late luteal	84	111 (74.4-163)	396 (373-422)	815 (703-908)

Test subjects	N	2.5th percentile pg/mL (90 % CI)	Median pg/mL (90 % CI)	97.5th percentile pg/mL (90 % CI)
Healthy men	150	11.3 (6.1-13.4)	24.8 (23.1-26.6)	43.2 (41.0-91.9)
Healthy postmenopausal women				
• Postmenopause	142	< 5 (< 5-< 5)	<5 (< 5-5.24)	138 (51.6-314)
Healthy pregnant women				
• 1st trimester	136	154 (127-173)	854 (737-1091)	3243 (2695-4161)
• 2nd trimester	140	1561 (1137-2032)	7739 (6596-8744)	21280 (18840-25130)
• 3rd trimester	136	8525 (7398-9312)	17625 (16990-18580)	> 30000 (29200-> 30000)

Healthy women Cycle Phase	N	5th percentile pg/mL (90 % CI)	Median pg/mL (90 % CI)	95th percentile pg/mL (90 % CI)
Follicular	85	30.9 (5.21-36.7)	53.9 (51.1-56.6)	90.4 (87.7-173)
Ovulation	81	60.4 (26.8-77)	206 (181-257)	533 (435-908)
Luteal	85	60.4 (43.2-76)	112 (106-133)	232 (207-363)

Healthy women Cycle Sub-Phase	N	5th percentile pg/mL (90 % CI)	Median pg/mL (90 % CI)	95th percentile pg/mL (90 % CI)
Early follicular	78	20.5 (< 5-21.4)	34 (32.6-36.7)	62.8 (52.1-77)
Intermediate follicular	83	26 (5.21-31)	46.9 (43.2-49)	79.8 (71.4-189)
Late follicular	84	49.5 (22.8-58.5)	126 (115-141)	233 (193-364)
Ovulation	79	60.4 (26.8-77)	222 (197-265)	602 (435-908)
Early luteal	85	51.1 (44.3-59.2)	106 (89.8-112)	179 (166-379)
Intermediate luteal	81	66.5 (42.7-90.7)	137 (121-155)	305 (256-418)
Late luteal	84	30.2 (20.2-44.3)	108 (101-115)	222 (191-247)

Each laboratory should investigate the transferability of the expected values to its own patient population and if necessary determine its own reference ranges.

### Specific performance data

Representative performance data on the analyzers are given below. Results obtained in individual laboratories may differ.

### Precision

Precision was determined using Elecsys reagents, samples and controls in a protocol (EP5-A2) of the CLSI (Clinical and Laboratory Standards Institute): 2 runs per day in duplicate each for 21 days (n = 84). The following results were obtained:

cobas e 411 analyzer					
Sample	Mean pmol/L	Repeatability		Intermediate precision	
		SD pmol/L	CV %	SD pmol/L	CV %
Human serum 1	93.3	7.91	8.5	11.1	11.9
Human serum 2	166	7.75	4.7	11.4	6.8
Human serum 3	605	18.8	3.1	20.4	3.4
Human serum 4	5021	97.7	1.9	125	2.5
Human serum 5	10760	253	2.4	297	2.8
PC U <sup>a</sup> 1	316	11.0	3.5	14.1	4.5
PC U2	1514	47.8	3.2	53.7	3.5

c) PC U = PreciControl Universal

cobas e 601 and cobas e 602 analyzers					
Sample	Mean pmol/L	Repeatability		Intermediate precision	
		SD pmol/L	CV %	SD pmol/L	CV %
Human serum 1	101	6.79	6.7	10.6	10.6
Human serum 2	173	6.46	3.7	9.8	5.7
Human serum 3	584	9.62	1.6	14.4	2.5
Human serum 4	4661	53.2	1.1	86.2	1.9
Human serum 5	9982	189	1.9	295	3.0
PC U1	329	7.82	2.4	12.4	3.8
PC U2	1497	18.0	1.2	30.8	2.1

cobas e 411 analyzer					
Sample	Mean pg/mL	Repeatability		Intermediate precision	
		SD pg/mL	CV %	SD pg/mL	CV %
Human serum 1	25.4	2.16	8.5	3.03	11.9
Human serum 2	45.3	2.11	4.7	3.10	6.8
Human serum 3	165	5.14	3.1	5.55	3.4
Human serum 4	1368	26.6	1.9	34.1	2.5
Human serum 5	2932	68.9	2.4	80.9	2.8
PC U1	86.1	2.99	3.5	3.83	4.5
PC U2	413	13.0	3.2	14.6	3.5

cobas e 601 and cobas e 602 analyzers					
Sample	Mean pg/mL	Repeatability		Intermediate precision	
		SD pg/mL	CV %	SD pg/mL	CV %
Human serum 1	27.4	1.85	6.7	2.90	10.6
Human serum 2	47.1	1.76	3.7	2.67	5.7
Human serum 3	159	2.62	1.6	3.92	2.5
Human serum 4	1270	14.5	1.1	23.5	1.9
Human serum 5	2720	51.5	1.9	80.5	3.0
PC U1	89.6	2.13	2.4	3.37	3.8
PC U2	408	4.91	1.2	8.39	2.1

### Method comparison

A comparison of the Elecsys Estradiol III assay (y) with ID-GC/MS (x) using clinical samples gave the following correlations (pg/mL):

Number of samples measured: 25

Passing/Bablok <sup>5</sup>	Linear regression
$y = 0.993x + 1.26$	$y = 1.00x + 2.07$
$r = 0.987$	$r = 0.999$

The sample concentrations were between 37.4 and 10768 pmol/L (10.2 and 2934 pg/mL) (ID-GC-MS concentrations).

### Analytical specificity

For the Elecsys Estradiol III assay, the following cross-reactivities were found (in %):

#### a) Substance added in a concentration of 0.001 µg/mL:

6-α-Hydroxy-Estradiol 74.1

#### b) Substance added in a concentration of 0.01 µg/mL:

4-Hydroxyestradiol 0.754

#### c) Substance added in a concentration of 0.1 µg/mL:

Aldosterone	0.005
Androstenedione	0.005
Equiline	0.057
Estriol	0.233
Estrone	0.757
Estrone-3β-glucuronide	0.003
Estrone-3-sulfate	0.002
Ethisterone	0.002
Norethindrone acetate	n. d. <sup>d)</sup>
Pregnenolone	0.007
Progesterone	0.004
2-Methoxyestradiol	0.121
17β-Estradiol-3,17-sulfate	0.002
17β-Estradiol-3-β-D-glucuronide	0.008
17β-Estradiol-17-β-D-glucuronide	0.001
17β-Estradiol-3-glucuronide-17-sulfate	n. d.
17β-Estradiol-3-sulfate-17-glucuronide	0.004
17β-Estradiol-3-sulfate	0.009
17β-Estradiol-17-valerate	0.163
17β-Estradiol-17-sulfate	0.003
2-Hydroxyestradiol	0.045

17-Hydroxyprogesterone n. d.

d) n. d. = not detectable

#### d) Substance added in a concentration of 0.2 µg/mL:

17-α-Ethinylestradiol	0.334
Cortisol	0.003
Cortisone	0.001
Tamoxifen	0.001

#### e) Substance added in a concentration of 0.25 µg/mL:

Chlomiphene 0.001

#### f) Substance added in a concentration of 1.0 µg/mL:

Prednisolone n. d.

#### g) Substance added in a concentration of 10 µg/mL:

Danazol	0.001
DHEA-S	n. d.
Mesterolone	n. d.
Testosterone	n. d.
5-α-Dihydrotestosterone (DHT)	n. d.
5-Androstene-3β-,17β-diol	n. d.

### References

- Johnson MR, Carter G, Grint C, et al. Relationship between ovarian steroids, gonadotropin and relaxin during the menstrual cycle. *Acta Endocrinol* 1993;129/2:121-125.
- Iqbal MJ, Dalton M, Sawers RS. Binding of testosterone and oestradiol to sex hormone binding globulin, human serum albumin and other plasma proteins: evidence for non-specific binding of oestradiol to sex hormone binding globulin. *Clin Science* 1983;64:307-314.
- Kronenberg HM, Melmed S, Polonsky KS, et al: *Williams Textbook of Endocrinology*. Saunders Elsevier 2008; Edition 11. ISBN 9781416029113.
- Thienpont LM, Verhseghe PG, Van Brussel KA, et al. Estradiol-17-beta quantified in serum by isotope dilution-gas chromatography-mass spectrometry. *Clin Chem* 1988(34);10:2066-2069.
- Bablok W, Passing H, Bender R, et al. A general regression procedure for method transformation. Application of linear regression procedures for method comparison studies in clinical chemistry, Part III. *J Clin Chem Clin Biochem* 1988 Nov;26(11):783-790.

For further information, please refer to the appropriate operator's manual for the analyzer concerned, the respective application sheets and the Method Sheets of all necessary components (if available in your country).

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the Member State in which the user and/or the patient is established.

### Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard (for USA: see dialog.roche.com for definition of symbols used):

	Contents of kit
	Analyzers/Instruments on which reagents can be used
	Reagent
	Calibrator
	Volume for reconstitution
	Global Trade Item Number

# Elecsys Estradiol III



## FOR US CUSTOMERS ONLY: LIMITED WARRANTY

Roche Diagnostics warrants that this product will meet the specifications stated in the labeling when used in accordance with such labeling and will be free from defects in material and workmanship until the expiration date printed on the label. THIS LIMITED WARRANTY IS IN LIEU OF ANY OTHER WARRANTY, EXPRESS OR IMPLIED, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR PARTICULAR PURPOSE. IN NO EVENT SHALL ROCHE DIAGNOSTICS BE LIABLE FOR INCIDENTAL, INDIRECT, SPECIAL OR CONSEQUENTIAL DAMAGES.

---

COBAS, COBAS E, ELECSYS and PRECICONTROL are trademarks of Roche. INTRALIPID is a trademark of Fresenius Kabi AB.

All other product names and trademarks are the property of their respective owners.

Additions, deletions or changes are indicated by a change bar in the margin.

© 2022, Roche Diagnostics



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim  
[www.roche.com](http://www.roche.com)

+800 5505 6606



Distribution in USA by:

Roche Diagnostics, Indianapolis, IN

US Customer Technical Support 1-800-428-2336