

Order information

REF	CONTENT	System-ID	Analyzer(s) on which cobas c pack(s) can be used
04490843 190	ONLINE TDM Gentamicin (100 tests)	07 6922 3	cobas c 311 , cobas c 501/502
Materials required (but not provided):			
03375790 190	Preciset TDM I Calibrators CAL A-F (1 x 5 mL) Preciset TDM I Calibrators Diluent (1 x 10 mL)	Codes 691-696	
04521536 190	TDM Control Set Level I (2 x 5 mL) TDM Control Set Level II (2 x 5 mL) TDM Control Set Level III (2 x 5 mL)	Code 310 Code 311 Code 312	

English**System information**For **cobas c 311/501** analyzer:**GENT2**: ACN 416For **cobas c 502** analyzer:**GENT2**: ACN 8416**Intended use**In vitro test for the quantitative determination of gentamicin in serum and plasma on Roche/Hitachi **cobas c** systems.**Summary**

Gentamicin is an aminoglycoside antibiotic that displays broad spectrum, high potency, anti-bacterial action for most susceptible organisms.^{1,2,3,4,5,6,7,8,9} At therapeutic serum concentrations ranging from 4 to 10 µg/mL (8.4 to 20.9 µmol/L), gentamicin is capable of inhibiting the growth of many gram positive cocci, especially penicillinase-producing staphylococci. At concentrations of 10 µg/mL (20.9 µmol/L), most strains of *E. coli*, *Proteus spp.*, *Klebsiella*, *Aerobacter*, *Clostridium*, *Brucella spp.*, *Salmonella*, *Serratia*, and *Shigella* are inhibited. At concentrations ranging from 4 to 10 µg/mL (8.4 to 20.9 µmol/L), gentamicin displays activity against most strains of *Pseudomonas aeruginosa*. Because of these characteristics, gentamicin has been most successfully used in the treatment of serious infections, especially those caused by gram-negative bacilli.^{10,11}

The therapeutic range of gentamicin should be measured at peak as well as trough concentrations. In patients with pre-existing renal damage or those to whom gentamicin has been administered for prolonged periods or in doses above the therapeutic range, hearing impairment and/or nephrotoxicity may develop. Therefore, monitoring of peak and trough gentamicin levels is critical in the prevention of these serious complications with the adjustment of dosage administration as indicated.^{12,13}

Test principle

The assay is based on the kinetic interaction of microparticles in a solution (KIMS). Gentamicin antibody is covalently coupled to microparticles and the drug derivative is linked to a macromolecule. The kinetic interaction of microparticles in solutions is induced by binding of drug-conjugate to the antibody on the microparticles and is inhibited by the presence of gentamicin in the sample. A competitive reaction takes place between the drug conjugate and gentamicin in the serum sample for binding to the gentamicin antibody on the microparticles. The resulting kinetic interaction of microparticles is indirectly proportional to the amount of drug present in the sample.

Reagents - working solutions

- R1** Gentamicin conjugate; piperazine-N,N'-bis (ethanesulfonic acid) (PIPES) buffer, pH 7.2; preservative
- R2** Anti-gentamicin antibody (mouse monoclonal); latex microparticle; 3-(N-morpholino) propane sulfonic acid (MOPS) buffer, pH 7.5; stabilizer; preservative

R1 is in position B and R2 is in position C.

Precautions and warnings

For in vitro diagnostic use for health care professionals. Exercise the normal precautions required for handling all laboratory reagents.

Infectious or microbial waste:

Warning: handle waste as potentially biohazardous material. Dispose of waste according to accepted laboratory instructions and procedures.

Environmental hazards:

Apply all relevant local disposal regulations to determine the safe disposal.

Safety data sheet available for professional user on request.

For USA: Caution: Federal law restricts this device to sale by or on the order of a physician.

Reagent handling

Ready for use

Carefully invert reagent container several times prior to use to ensure that the reagent components are mixed.

Storage and stability

Shelf life at 2-8 °C:

See expiration date on **cobas c** pack label

On-board in use and refrigerated on the analyzer:

12 weeks

Do not freeze.**Specimen collection and preparation**

For specimen collection and preparation only use suitable tubes or collection containers.

Only the specimens listed below were tested and found acceptable.

Serum: Collect serum using standard sampling tubes.

Plasma: K₂- or K₃-EDTA, sodium citrate, or sodium, lithium, or ammonium heparin plasma.

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube manufacturer.

Stability:

1 week capped at 2-8 °C

4 weeks capped at -20 °C

Centrifuge samples containing precipitates before performing the assay.

See the limitations and interferences section for details about possible sample interferences.

Sample stability claims were established by experimental data by the manufacturer or based on reference literature and only for the temperatures/time frames as stated in the method sheet. It is the responsibility of the individual laboratory to use all available references and/or its own studies to determine specific stability criteria for its laboratory.

Specimens should not be repeatedly frozen and thawed.

Invert thawed specimens several times prior to testing.

Usual sampling time varies dependent upon desired measurement of peak or through values.¹⁴**Materials provided**

See "Reagents – working solutions" section for reagents.

Materials required (but not provided)

- See "Order information" section
- General laboratory equipment

Assay

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

The performance of applications not validated by Roche is not warranted and must be defined by the user.

Application for serum and plasma

Deselect Automatic Rerun for these applications in the Utility menu, Application screen, Range tab.

cobas c 311 test definition

Assay type	2-Point End		
Reaction time / Assay points	10 / 11-32		
Wavelength (sub/main)	800/600 nm		
Reaction direction	Increase		
Unit	µg/mL (µmol/L)		
Reagent pipetting		Diluent (H ₂ O)	
R1	100 µL	-	
R2	95 µL	-	
Sample volumes	Sample	Sample dilution	
		Sample	Diluent (H ₂ O)
Normal	2.0 µL	-	-
Decreased	2.0 µL	-	-
Increased	2.0 µL	-	-

cobas c 501/502 test definition

Assay type	2-Point End		
Reaction time / Assay points	10 / 18-49		
Wavelength (sub/main)	800/600 nm		
Reaction direction	Increase		
Unit	µg/mL (µmol/L)		
Reagent pipetting		Diluent (H ₂ O)	
R1	100 µL	-	
R2	95 µL	-	
Sample volumes	Sample	Sample dilution	
		Sample	Diluent (H ₂ O)
Normal	2.0 µL	-	-
Decreased	2.0 µL	-	-
Increased	2.0 µL	-	-

Calibration

Calibrators	S1-6: Preciset TDM I calibrators
Calibration mode	RCM
Calibration frequency	6-point calibration - after cobas c pack change - after reagent lot change - as required following quality control procedures

ACTION REQUIRED

After any calibration with Preciset TDM I Calibrators, the TDM Control Set have to be run in the order Level 3 to Level 1. Prior to running samples run a blank serum sample. The blank serum sample can be scheduled for any R1/R2 assay.

Calibration interval may be extended based on acceptable verification of calibration by the laboratory.

Traceability: This method has been standardized against USP reference standards. The calibrators are prepared to contain known quantities of gentamicin in normal human serum.

Quality control

For quality control, use control materials as listed in the "Order information" section.

In addition, other suitable control material can be used.

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

Follow the applicable government regulations and local guidelines for quality control.

Calculation

cobas c systems automatically calculate the analyte concentration of each sample.

Conversion factor:¹⁵ µg/mL x 2.09 = µmol/L

Limitations - interference

Criterion: Recovery within ± 10 % of initial value at gentamicin levels of approximately 2 and 6 µg/mL (4.2 and 12.5 µmol/L).

Serum/Plasma

Icterus:¹⁶ No significant interference up to an I index of 50 for conjugated and unconjugated bilirubin (approximate conjugated and unconjugated bilirubin concentration: 855 µmol/L or 50 mg/dL).

Hemolysis:¹⁶ No significant interference up to an H index of 1000 (approximate hemoglobin concentration: 621 µmol/L or 1000 mg/dL).

Lipemia (Intralipid):¹⁶ No significant interference up to an L index of 150. There is poor correlation between the L index (corresponds to turbidity) and triglycerides concentration.

Triglycerides: No significant interference from triglycerides up to a concentration of 1000 mg/dL (11.3 mmol/L).

Rheumatoid factors: No significant interference from rheumatoid factors up to a concentration of 100 IU/mL.

Total protein: No significant interference from total protein up to a concentration of 12 g/dL.

Note

A negative bias of up to approximately 20 % has been observed with this assay for some samples artificially spiked with Gentamicin sulfate. Patient samples have been verified to recover correctly.

For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.

ACTION REQUIRED

Special Wash Programming: The use of special wash steps is mandatory when certain test combinations are run together on **cobas c** systems. The latest version of the carry-over evasion list can be found with the NaOH-SMS-SmpCln1+2-SCCS Method Sheets. For further instructions refer to the operator's manual. **cobas c 502** analyzer: All special wash programming necessary for avoiding carry-over is available via the **cobas** link, manual input is required in certain cases.

Where required, special wash/carry-over evasion programming must be implemented prior to reporting results with this test.

Limits and ranges**Measuring range**

Measuring range: 0.4-10.0 µg/mL (0.84-20.9 µmol/L)

Manually dilute samples having higher concentrations with Preciset TDM I diluent (0 µg/mL) (1 + 1) and reassay. Multiply the result by 2 to obtain the specimen value.

Lower limits of measurement

Lower detection limit of the test

0.3 µg/mL (0.63 µmol/L)

The lower detection limit represents the lowest measurable analyte level that can be distinguished from zero. It is calculated as the value lying

2 standard deviations above that of the 0 µg/mL calibrator (standard 1 + 2 SD, repeatability, n = 21).

Functional sensitivity

0.4 µg/mL (0.84 µmol/L)

The functional sensitivity is the lowest analyte concentration that can be reproducibly measured with a coefficient of variation of ≤ 20 % (repeatability, n = 21).

Expected values

Although optimum values may vary, peak serum values in the range of 6 to 10 µg/mL (12.5 to 20.9 µmol/L) and trough values in the range of 0.5 to 2.0 µg/mL (1.0 to 4.2 µmol/L) are generally accepted for therapeutic effectiveness.¹⁷ The achievement of non-toxic, but therapeutic, serum levels is often difficult, even in patients with normal renal function. Complications encountered with the use of gentamicin are ototoxicity and nephrotoxicity.^{10,18,19,20,21} However, these reactions are predictable, and close patient monitoring is essential for the successful use of this agent. The most serious toxic effect of gentamicin is permanent damage to the vestibular division of the eighth cranial nerve, which has been reported to occur most frequently in patients with renal failure. Since gentamicin is inherently unstable, is not metabolized and is excreted primarily by glomerular filtration, toxic concentrations of the drug may accumulate in the body when the dosage is not adjusted for patients with impaired renal function. While high serum levels can be toxic, indiscriminately low dosages of gentamicin will result in ineffective treatment for many strains of gram-negative bacteria. The indiscriminate use of low dosages of gentamicin may not only engender the emergence of gentamicin-resistant organisms, but also the emergence of aminoglycoside-resistant organisms.^{11,22,23} Current literature reflects increasing interest in once daily dosing versus the conventional administration of drug 2 to 4 times daily. Adoption of once daily dosing may require a revision of target peak and trough concentrations.^{24,25,26}

Expected values reflect the data and information provided in the reference and do not represent therapeutic recommendations and / or dosage instructions. For therapeutic recommendations and dosage instructions refer to the respective national guidelines and the full prescription information of the drug.

Each laboratory should investigate the transferability of the expected values to its own patient population and if necessary determine its own reference ranges.

Specific performance data

Representative performance data on the analyzers are given below. Results obtained in individual laboratories may differ.

Precision

Precision was determined using human samples and controls in a modified NCCLS EP5-T2 protocol (repeatability n = 63, intermediate precision n = 63). The following results were obtained on a **cobas c 501** analyzer.

Serum/Plasma

Repeatability	Mean		SD		CV
	µg/mL	µmol/L	µg/mL	µmol/L	%
Control 1	1.87	3.91	0.08	0.17	4.2
Control 2	4.37	9.13	0.08	0.17	1.8
Control 3	6.48	13.5	0.12	0.3	1.8
HS 1	1.90	3.97	0.07	0.15	3.6
HS 2	6.04	12.6	0.13	0.3	2.2

Intermediate precision	Mean		SD		CV
	µg/mL	µmol/L	µg/mL	µmol/L	%
Control 1	1.87	3.91	0.09	0.19	5.1
Control 2	4.37	9.13	0.09	0.19	2.0
Control 3	6.48	13.5	0.15	0.31	2.3
HS 1	1.90	3.97	0.08	0.17	4.1
HS 2	6.04	12.6	0.17	0.36	2.8

The data obtained on **cobas c 501** analyzer(s) are representative for **cobas c 311** analyzer(s).

Method comparison**Serum/Plasma**

Gentamicin values for human serum and plasma samples obtained on a **cobas c 501** analyzer (y) were compared with those determined using the corresponding reagent on a Roche/Hitachi 917 analyzer (x) and on a COBAS INTEGRA 800 analyzer (x).

Roche/Hitachi 917 analyzer

Sample size (n) = 64

Passing/Bablok²⁷

Linear regression

$y = 0.976x - 0.014 \mu\text{g/mL}$

$y = 0.976x - 0.016 \mu\text{g/mL}$

$r = 0.962$

$r = 0.999$

The sample concentrations were between 0.540 and 9.65 µg/mL (1.13 and 20.2 µmol/L).

COBAS INTEGRA 800 analyzer

Sample size (n) = 63

Passing/Bablok²⁷

Linear regression

$y = 0.983x - 0.214 \mu\text{g/mL}$

$y = 0.988x - 0.219 \mu\text{g/mL}$

$r = 0.965$

$r = 0.997$

The sample concentrations were between 0.530 and 9.14 µg/mL (1.11 and 19.1 µmol/L).

The data obtained on **cobas c 501** analyzer(s) are representative for **cobas c 311** analyzer(s).

Analytical specificity

The following compounds were tested for cross-reactivity.

Compound	Concentration	%
	Tested (µg/mL)	Cross-Reactivity
Netilmicin	70	9.13
Sisomicin	131	8.16
Methotrexate	23	< 1.0
Tetracycline	40	< 1.0
Amikacin	250	< 0.1
Cephalexin	500	< 0.1
Chloramphenicol	300	< 0.1
Clindamycin	500	< 0.1
Kanamycin	250	< 0.1
Neomycin	100	< 0.1
Spectinomycin	200	< 0.1
Streptomycin	200	< 0.1
Tobramycin	100	< 0.1
Vancomycin	400	< 0.1
Amphotericin B	50	< 0.01
Ampicillin	78	< 0.01
Carbenicillin	500	< 0.01
Cephalosporin C	432	< 0.01
Cephalothin	63	< 0.01
Erythromycin	200	< 0.01
5-Fluorocytosine	700	< 0.01
Furosemide	100	< 0.01
Methylprednisolone	500	< 0.01
Oxytetracycline	37	< 0.01
Prednisolone	500	< 0.01

Tests were performed on 16 drugs. No significant interference with the assay was found.

Acetaminophen	Doxycycline (Tetracycline)
Acetyl cysteine	Ibuprofen
Acetylsalicylic acid	Levodopa
Ampicillin-Na	Methyldopa + 1.5 H ₂ O
Ascorbic acid	Metronidazole
Ca-Dobesilate	Phenylbutazone
Cefoxitin	Rifampicin
Cyclosporine	Theophylline

References

- Black J, Calesnick B, Williams D, et al. Pharmacology of gentamicin, a new broad-spectrum antibiotic. *Antimicrob Agents Chemother* 1963;161:138-147.
- Oden EM, Stander H, Weinstein MJ. Microbiological assay of gentamicin. *Antimicrob Agents Chemother* 1963;161:8-13.
- Waitz JA, Weinstein M. Recent laboratory studies with gentamicin with reference to structure and resistance. *Med J Aust* 1970;1(Suppl):19-25.
- MacMillan BG. Ecology of bacteria colonizing the burned patient given topical and systemic gentamicin therapy: a five-year study. *J Infect Dis* 1971;124(Suppl):278-286.
- Nunnery AW, Riley HD Jr. Gentamicin: clinical and laboratory studies in infants and children. *J Infect Dis* 1969;119(4):460-464.
- Young LS, Luria DB, Armstrong D. Gentamicin in the treatment of severe, hospital-acquired gram-negative infections. *Trans NY Acad Sci* 1967;29(5):579-588.
- Holloway WJ, Taylor WA. Gentamicin and kanamycin in the treatment of gram-negative sepsis: a comparative study. *J Infect Dis* 1971;124(Suppl):180-184.
- Bodey GP, Middleman E, Umsawadi T, et al. Intravenous gentamicin therapy for infections in patients with cancer. *J Infect Dis* 1971;124(Suppl):174-179.
- Klein JO, Herschel M, Therakan RM, et al. Gentamicin in serious neonatal infections: absorption, excretion, and clinical results in 25 cases. *J Infect Dis* 1971;124(Suppl):224-231.
- Cox CE. Gentamicin. *Med Clin North Am* 1970;54:1305-1315.
- Riff LJ, Jackson GG. Pharmacology of gentamicin in man. *J Infect Dis* 1971;124(Suppl):98-105.
- Kahlmeter G. Gentamicin and tobramycin. Clinical pharmacokinetics and nephrotoxicity. Aspects on assay techniques. *Scand J Infect Dis* 1979;132(Suppl 18):1-40.
- Barza M, Lauerhmann M. Why monitor serum levels of gentamicin? *Clin Pharm* 1978;3:202-215.
- O'Bey KA, Jim LK, Gee JP, et al. Temperature dependence of the stability of tobramycin mixed with penicillins in human serum. *Am J Hosp Pharm* 1982;39(6):1005-1008.
- Tietz NW. *Fundamentals of Clinical Chemistry*. 3rd ed. Philadelphia, PA: WB Saunders Co 1995:834.
- Glick MR, Ryder KW, Jackson SA. Graphical Comparisons of Interferences in Clinical Chemistry Instrumentation. *Clin Chem* 1986;32:470-475.
- Taylor WJ, Diers-Caviness MH (eds.). *A Textbook for the Clinical Application of Therapeutic Drug Monitoring*. Abbott Laboratories 1986:296.
- Jackson GG, Arcieri GM. Ototoxicity of gentamicin in man: a survey and controlled analysis of clinical experience in the United States. *J Infect Dis* 1971;124(Suppl):130-137.
- Arcieri GM, Falco FG, Smith HM, et al. Clinical research experience with gentamicin. Incidence of adverse reactions. *Med J Aust* 1970;1(Suppl):30-34.
- Falco FG, Smith HM, Arcieri GM. Nephrotoxicity of aminoglycosides and gentamicin. *J Infect Dis* 1969;119(4):406-409.
- Gyselynck AM, Forrey A, Cutler R. Pharmacokinetics of gentamicin: distribution and plasma and renal clearance. *J Infect Dis* 1971;124(Suppl):70-76.
- Weinstein MJ, Drube CG, Moss EL Jr, et al. Microbiologic studies related to bacterial resistance to gentamicin. *J Infect Dis* 1971;124(Suppl):11-17.
- Jackson GG, Riff LJ. Pseudomonas bacteremia: pharmacologic and other bases for failure of treatment with gentamicin. *J Infect Dis* 1971;124(Suppl):185-191.
- Barclay ML, Begg EJ, Hickling KG. What is the evidence for once-daily aminoglycoside therapy? *Clin Pharmacokinet* 1994;27(1):32-48.
- Hustinx WN, Hoepelman IM. Aminoglycoside dosage regimens. Is once a day enough? *Clin Pharmacokinet* 1993;25(6):427-432.
- Bertino JS Jr, Rodvold KA, Destache CJ. Cost considerations in therapeutic drug monitoring of aminoglycosides. *Clin Pharmacokinet* 1994;26(1):71-81.
- Bablok W, Passing H, Bender R, et al. A general regression procedure for method transformation. Application of linear regression procedures for method comparison studies in clinical chemistry, Part III. *J Clin Chem Clin Biochem* 1988 Nov;26(11):783-790.

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the Member State in which the user and/or the patient is established.

The Summary of Safety & Performance Report can be found here: <https://ec.europa.eu/tools/eudamed>

Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard (for USA: see dialog.roche.com for definition of symbols used):

 CONTENT

Contents of kit



Volume after reconstitution or mixing

 GTIN

Global Trade Item Number

FOR US CUSTOMERS ONLY: LIMITED WARRANTY

Roche Diagnostics warrants that this product will meet the specifications stated in the labeling when used in accordance with such labeling and will be free from defects in material and workmanship until the expiration date printed on the label. **THIS LIMITED WARRANTY IS IN LIEU OF ANY OTHER WARRANTY, EXPRESS OR IMPLIED, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR PARTICULAR PURPOSE. IN NO EVENT SHALL ROCHE DIAGNOSTICS BE LIABLE FOR INCIDENTAL, INDIRECT, SPECIAL OR CONSEQUENTIAL DAMAGES.**

COBAS, COBAS C, COBAS INTEGRA, ONLINE TDM and PRECISET are trademarks of Roche.

All other product names and trademarks are the property of their respective owners.

Additions, deletions or changes are indicated by a change bar in the margin.

© 2021, Roche Diagnostics



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim
www.roche.com

+800 5505 6606



Distribution in USA by:
Roche Diagnostics, Indianapolis, IN
US Customer Technical Support 1-800-428-2336