

**Rheumatoid Factors II****Order information**

REF	CONTENT	Analyzer(s) on which <b>cobas c</b> pack(s) can be used
20764574 322	Rheumatoid Factors II 100 tests	System-ID 07 6457 4 <b>cobas c 311, cobas c 501/502</b>
12172828 322	Preciset RF (5 x 1 mL)	Codes 725-729
03005496 122	RF Control Set (4 x 1 mL)	Code 215 Level I
		Code 216 Level II
04489357 190	Diluent NaCl 9 % (50 mL)	System-ID 07 6869 3

**English****System information**

For **cobas c** 311/501 analyzers:

**RF-II:** ACN 017

For **cobas c** 502 analyzer:

**RF-II:** ACN 8017

**Intended use**

In vitro test for the quantitative determination of Rheumatoid Factors (RF-II) in human serum and plasma on Roche/Hitachi **cobas c** systems. Measurements may be used as an aid in the diagnosis of rheumatoid arthritis.

**Summary**<sup>1,2,3,4,5,6,7,8,9,10</sup>

Rheumatoid factors are a heterogeneous group of autoantibodies directed against the antigenic determinants on the Fc-region of IgG molecules. They are important in the diagnosis of rheumatoid arthritis, but can also be found in other inflammatory rheumatic diseases and in various non-rheumatic diseases. They are also found in clinically healthy persons over 60 years of age. Despite these restrictions, the detection of rheumatoid factors is a diagnostic criterion of the American College of Rheumatology for classifying rheumatoid arthritis. The autoantibodies occur in all the immunoglobulin classes, although the usual analytical methods are limited to the detection of rheumatoid factors of the IgM type.

The classic procedure for the quantitation of rheumatoid factors is by agglutination with IgG-sensitized sheep erythrocytes or latex particles. Particular problems of these semiquantitative methods are the poor between-laboratory precision and reproducibility, together with standardization difficulties. For these reasons, new assay methods such as nephelometry, turbidimetry, enzyme-immunoassays and radioimmunoassays have been developed. The Roche RF assay is based on the immunological agglutination principle with enhancement of the reaction by latex.

**Test principle**<sup>4,5,6</sup>

Immunoturbidimetric assay.

Latex-bound heat-inactivated IgG (antigen) reacts with the RF-antibodies in the sample to form antigen/antibody complexes which, following agglutination, are measured turbidimetrically.

**Reagents - working solutions**

**R1** Glycine buffer: 170 mmol/L, pH 8.0; polyethylene glycol: 0.05 %; bovine serum albumin; stabilizer; preservative

**R2** Latex particles coated with human IgG; glycine buffer: 170 mmol/L, pH 7.3; stabilizer; preservative

R1 is in position B and R2 is in position C.

**Precautions and warnings**

For in vitro diagnostic use.

Exercise the normal precautions required for handling all laboratory reagents.

Disposal of all waste material should be in accordance with local guidelines. Safety data sheet available for professional user on request.

All human material should be considered potentially infectious. All products derived from human blood are prepared exclusively from the blood of donors tested individually and shown to be free from HBsAg and antibodies to HCV and HIV. The testing methods used assays approved by the FDA or cleared in compliance with the European Directive 98/79/EC, Annex II, List A.

However, as no testing method can rule out the potential risk of infection with absolute certainty, the material should be handled with the same level

of care as a patient specimen. In the event of exposure, the directives of the responsible health authorities should be followed.<sup>11,12</sup>

For USA: Caution: Federal law restricts this device to sale by or on the order of a physician.

This kit contains components classified as follows in accordance with the Regulation (EC) No. 1272/2008:

2-methyl-2H-isothiazol-3-one hydrochloride

EUH 208 May produce an allergic reaction.

Product safety labeling follows EU GHS guidance.

**Reagent handling**

Ready for use

Mix **cobas c** pack well before placing on the analyzer.

Carefully invert reagent container several times prior to use to ensure that the reagent components are mixed.

**Storage and stability****RF-II**

Shelf life at 2-8 °C: See expiration date on **cobas c** pack label.

On-board in use and refrigerated on the analyzer: 8 weeks

**Diluent NaCl 9 %**

Shelf life at 2-8 °C: See expiration date on **cobas c** pack label.

On-board in use and refrigerated on the analyzer: 12 weeks

**Specimen collection and preparation**

For specimen collection and preparation only use suitable tubes or collection containers.

Only the specimens listed below were tested and found acceptable.

Serum

Plasma: Li-heparin and K<sub>2</sub>-EDTA plasma

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube manufacturer.

Centrifuge samples containing precipitates before performing the assay.

See the limitations and interferences section for details about possible sample interferences.

Stability:<sup>13</sup> 1 day at 20-25 °C  
8 days at 4-8 °C  
3 months at -20 °C (freeze only once)

Sample stability claims were established by experimental data by the manufacturer or based on reference literature and only for the temperatures/time frames as stated in the method sheet. It is the responsibility of the individual laboratory to use all available references and/or its own studies to determine specific stability criteria for its laboratory.

## Rheumatoid Factors II

**Materials provided**

See "Reagents – working solutions" section for reagents.

**Materials required (but not provided)**

- See "Order information" section
- General laboratory equipment

**Assay**

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

The performance of applications not validated by Roche is not warranted and must be defined by the user.

**Application for serum and plasma****cobas c 311 test definition**

Assay type	2-Point End		
Reaction time/Assay points	10 / 7-18		
Wavelength (sub/main)	800/570 nm		
Reaction direction	Increase		
Unit	IU/mL		
Reagent pipetting	Diluent (H <sub>2</sub> O)		
R1	90 µL	–	
R2	30 µL	–	
<i>Sample volumes</i>	<i>Sample</i>	<i>Sample dilution</i>	
		<i>Sample</i>	<i>Diluent (NaCl)</i>
Normal	3 µL	–	–
Decreased	6 µL	15 µL	135 µL
Increased	3 µL	–	–

**cobas c 501 test definition**

Assay type	2-Point End		
Reaction time/Assay points	10 / 12-26		
Wavelength (sub/main)	800/570 nm		
Reaction direction	Increase		
Unit	IU/mL		
Reagent pipetting	Diluent (H <sub>2</sub> O)		
R1	90 µL	–	
R2	30 µL	–	
<i>Sample volumes</i>	<i>Sample</i>	<i>Sample dilution</i>	
		<i>Sample</i>	<i>Diluent (NaCl)</i>
Normal	3 µL	–	–
Decreased	6 µL	15 µL	135 µL
Increased	3 µL	–	–

**cobas c 502 test definition**

Assay type	2-Point End		
Reaction time/Assay points	10 / 12-26		
Wavelength (sub/main)	800/570 nm		
Reaction direction	Increase		
Unit	IU/mL		
Reagent pipetting	Diluent (H <sub>2</sub> O)		
R1	90 µL	–	
R2	30 µL	–	
<i>Sample volumes</i>	<i>Sample</i>	<i>Sample dilution</i>	

		<i>Sample</i>	<i>Diluent (NaCl)</i>
Normal	3 µL	–	–
Decreased	6 µL	15 µL	135 µL
Increased	6 µL	–	–

**Calibration**

Calibrators	S1: H <sub>2</sub> O S2-6: Preciset RF
Calibration mode	RCM
Calibration frequency	Full calibration <ul style="list-style-type: none"> <li>• after 180 days during shelf life</li> <li>• after reagent lot change</li> <li>• as required following quality control procedures</li> </ul>

Calibration interval may be extended based on acceptable verification of calibration by the laboratory.

Traceability:<sup>14</sup> This method has been standardized using the WHO Standard 64/2.

**Quality control**

For quality control, use control materials as listed in the "Order information" section.

In addition, other suitable control material can be used.

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

Follow the applicable government regulations and local guidelines for quality control.

**Calculation**

**cobas c** systems automatically calculate the analyte concentration of each sample.

**Limitations - interference**

Criterion: Recovery within  $\pm 10\%$  of initial value at an RF concentration of 14 IU/mL.

Icterus:<sup>15</sup> No significant interference up to an I index of 40 for conjugated and 60 for unconjugated bilirubin (approximate conjugated bilirubin concentration: 624 µmol/L or 40 mg/dL) and approximate unconjugated bilirubin concentration: 1026 µmol/L or 60 mg/dL).

Hemolysis:<sup>15</sup> No significant interference up to an H index of 300 (approximate hemoglobin concentration: 186 µmol/L or 300 mg/dL).

Lipemia (Intralipid):<sup>15</sup> No significant interference up to an L index of 2000. There is poor correlation between the L index (corresponds to turbidity) and triglycerides concentration.

Drugs: No interference was found at therapeutic concentrations using common drug panels.<sup>16,17</sup>

High dose hook-effect: Using the prozone check, no false result without a flag was observed up to an RF concentration of 6000 IU/mL.

In very rare cases, gammopathy, in particular type IgM (Waldenström's macroglobulinemia), may cause unreliable results.<sup>18</sup>

There is the possibility that other substances and/or factors may interfere with the test and cause unreliable results.

For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.

**ACTION REQUIRED**

**Special Wash Programming:** The use of special wash steps is mandatory when certain test combinations are run together on **cobas c** systems. The latest version of the carry-over evasion list can be found with the NaOHD-SMS-SmpCln1+2-SCCS Method Sheets. For further instructions refer to the operator's manual. **cobas c 502** analyzer: All special wash programming necessary for avoiding carry-over is available via the **cobas** link, manual input is required in certain cases.

**Where required, special wash/carry-over evasion programming must be implemented prior to reporting results with this test.**

**Rheumatoid Factors II****Limits and ranges****Measuring range**

10-130 IU/mL

Determine samples having higher concentrations via the rerun function. Dilution of samples via the rerun function is a 1:5 dilution. Results from samples diluted using the rerun function are automatically multiplied by a factor of 5.

**Lower limits of measurement***Lower detection limit of the test*

10 IU/mL

The lower detection limit represents the lowest measurable analyte level that can be distinguished from zero. It is calculated as the value lying 3 standard deviations above that of the lowest standard (standard 1 + 3 SD, repeatability, n = 21).

**Expected values**

&lt; 14 IU/mL

This value is based on serum samples from 541 test subjects.

Each laboratory should investigate the transferability of the expected values to its own patient population and if necessary determine its own reference ranges.

**Specific performance data**

Representative performance data on the analyzers are given below. Results obtained in individual laboratories may differ.

**Precision**

Precision was determined using human samples and controls in an internal protocol with repeatability (n = 21) and intermediate precision (3 aliquots per run, 1 run per day, 21 days). The following results were obtained:

<i>Repeatability</i>	<i>Mean</i>	<i>SD</i>	<i>CV</i>
	<i>IU/mL</i>	<i>IU/mL</i>	<i>%</i>
RF Control level 1	23.7	0.2	0.8
RF Control level 2	53.0	0.5	0.9
Human serum 1	19.5	0.3	1.6
Human serum 2	27.5	0.3	1.1
<i>Intermediate precision</i>	<i>Mean</i>	<i>SD</i>	<i>CV</i>
	<i>IU/mL</i>	<i>IU/mL</i>	<i>%</i>
RF Control level 1	23.2	0.3	1.4
RF Control level 2	51.4	0.8	1.5
Human serum 3	19.3	0.3	1.6
Human serum 4	26.1	0.5	1.8

**Method comparison**

RF values for human serum and plasma samples obtained on a **cobas c** 501 analyzer (y) were compared with those determined using the corresponding reagent on a Roche/Hitachi 917 analyzer (x).

Sample size (n) = 70

Passing/Bablok <sup>19</sup>	Linear regression
y = 1.000x - 1.20 IU/mL	y = 0.999x - 1.39 IU/mL
τ = 0.959	r = 0.998

The sample concentrations were between 10.8 and 114 IU/mL.

**References**

- Borque L, Barozzi D, Ferrari L, et al. The Determination of Rheumatoid Factors by an Immunoturbidimetric Assay on Boehringer Mannheim/Hitachi Analysis Systems. *Klin Lab* 1994;40:445-453.
- Arnett FC, Edworthy SM, Bloch DA, et al. The American Rheumatism Association 1987 revised criteria for the classification of rheumatoid arthritis. *Arthritis Rheum* 1988;31:315-324.
- Bartfield H. Distribution of rheumatoid factor activity in non-rheumatoid states. *Ann NY Acad Sci* 1969;16:30-40.
- Waalder E. On the occurrence of a factor in human serum activating the specific agglutination of sheep blood corpuscles. *Acta Pathol Microbiol Scand* 1940;17:172-178.
- Moore TL, Dorner RN. Rheumatoid factors. *Clin Biochem* 1993;26:75-84.
- Roberts-Thomson PJ, McEvoy R, et al. *Ann Rheum Dis* 1985;44:379-383.
- Borque L, Yago M, Mar C, et al. *Clin Chem* 1986;32:124-129.
- Bampton JL, Cawston TE, Kyle MV, et al. *Ann Rheum Dis* 1985;44:379-383.
- Koopman WJ, Schrohenloker RE. *Arthritis Rheum* 1980;23:302-308.
- Jaspers JP, Van Oers RJM, Leerkes B. Nine Rheumatoid Factor Assays Compared. *J Clin Chem Clin Biochem* 1988;26:863-871.
- Occupational Safety and Health Standards: Bloodborne pathogens. (29 CFR Part 1910.1030). *Fed. Register*.
- Directive 2000/54/EC of the European Parliament and Council of 18 September 2000 on the protection of workers from risks related to exposure to biological agents at work.
- Use of Anticoagulants in Diagnostic Laboratory Investigations. WHO Publication WHO/DIL/LAB/99.1 Rev. 2: Jan 2002.
- Anderson SG, Bentzon MW, Houba V, et al. International reference preparation of rheumatoid arthritis serum. *Bull Wld Hlth Org* 1970;42:311-318.
- Glick MR, Ryder KW, Jackson SA. Graphical Comparisons of Interferences in Clinical Chemistry Instrumentation. *Clin Chem* 1986;32:470-475.
- Breuer J. Report on the Symposium "Drug effects in Clinical Chemistry Methods". *Eur J Clin Chem Clin Biochem* 1996;34:385-386.
- Sonntag O, Scholer A. Drug interference in clinical chemistry: recommendation of drugs and their concentrations to be used in drug interference studies. *Ann Clin Biochem* 2001;38:376-385.
- Bakker AJ, Mücke M. Gammopathy interference in clinical chemistry assays: mechanisms, detection and prevention. *Clin Chem Lab Med* 2007;45(9):1240-1243.
- Bablok W, Passing H, Bender R, et al. A general regression procedure for method transformation. Application of linear regression procedures for method comparison studies in clinical chemistry, Part III. *J Clin Chem Clin Biochem* 1988 Nov;26(11):783-790.

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

**Symbols**

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard (for USA: see dialog.roche.com for definition of symbols used):

CONTENT	Contents of kit
→	Volume after reconstitution or mixing
GTIN	Global Trade Item Number

**FOR US CUSTOMERS ONLY: LIMITED WARRANTY**

Roche Diagnostics warrants that this product will meet the specifications stated in the labeling when used in accordance with such labeling and will be free from defects in material and workmanship until the expiration date printed on the label. THIS LIMITED WARRANTY IS IN LIEU OF ANY OTHER WARRANTY, EXPRESS OR IMPLIED, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR PARTICULAR PURPOSE. IN NO EVENT SHALL ROCHE DIAGNOSTICS BE LIABLE FOR INCIDENTAL, INDIRECT, SPECIAL OR CONSEQUENTIAL DAMAGES.

COBAS, COBAS C and PRECISET are trademarks of Roche.

All other product names and trademarks are the property of their respective owners.

Additions, deletions or changes are indicated by a change bar in the margin.

© 2019, Roche Diagnostics

0020764574322c501V9.0

# RF-II

Rheumatoid Factors II



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim  
[www.roche.com](http://www.roche.com)

Distribution in USA by:  
Roche Diagnostics, Indianapolis, IN  
US Customer Technical Support 1-800-428-2336



# cobas®