

REF



SYSTEM

03052001190

03052001500

100

cobas e 411  
cobas e 601  
cobas e 602

## English

## System information

For **cobas e 411** analyzer: test number 750For **cobas e 601** and **cobas e 602** analyzers: Application Code Number 146

## Intended use

Immunoassay for the in vitro quantitative determination of sex hormone-binding globulin in human serum and plasma.

The electrochemiluminescence immunoassay "ECLIA" is intended for use on Elecsys and **cobas e** immunoassay analyzers.

## Summary

Sex hormone-binding globulin (SHBG) is the blood transport protein for testosterone and estradiol (E2). It is a large glycoprotein with a molecular weight of about 95 kDa, and exists as a homodimer composed of two identical subunits. Each subunit contains two disulfide bridges.<sup>1,2</sup> Generally, optimal steroid binding to SHBG will require a planar C19 steroid with a 17 $\alpha$ -OH group, and an electronegative functional group at C3, whereas dissociated subunits of SHBG do not bind steroid and the two subunits associate to form a single steroid-binding site.<sup>3</sup>

The level of SHBG binding affinity for steroids from high to low is as following: dihydrotestosterone, testosterone, androstenediol, E2 and estrone. SHBG also weakly binds with dehydroepiandrosterone (DHEA) however not with dehydroepiandrosterone sulfate (DHEA-S).<sup>4</sup>

SHBG is produced mainly by the liver and its synthesis and secretion are regulated by estrogen and negatively influenced by liver fat content and inflammatory cytokines.<sup>4,5,6,7,8,9</sup> Decreased SHBG serum levels are associated with conditions where elevated androgen levels are present or where the effect of androgen on its target organs is excessive. This explains the gender-related differences seen between men and women, especially during puberty. However, decreased SHBG levels are also seen in inflammation and in case of a diet leading to fat build-up in the liver e.g. rich in monosaccharides, particularly fructose. Low serum SHBG concentrations may correlate with cardiovascular disease risk,<sup>10,11</sup> type 2 diabetes,<sup>12,13</sup> as well as breast cancer.<sup>14</sup> High serum SHBG concentrations have been proposed to be associated with change of diet resulting in loss of weight. Moreover, there is increasing evidence that plasma SHBG levels could serve as a biomarker for diseases with chronic inflammation and as indicator of response to anti-inflammatory treatment.<sup>6</sup>

Low SHBG titer can be an important indicator of an excessive/chronic androgenic action where androgen levels are normal, but where clinical symptoms would seem to indicate androgen in excess.<sup>15</sup>

By calculating the free androgen index (FAI), also called free testosterone index (FTI), from the ratio of total testosterone (T) to SHBG [% FAI or FTI = (100 T/SHBG)], it is possible to calculate the approximate amount of free testosterone, as there is a direct correlation between FAI and FT. By additionally taking the non-specifically albumin-bound testosterone into account, it is possible to calculate the bioavailable testosterone, which is the sum of free testosterone and the albumin-bound testosterone fraction, calculated via the association constant to albumin.<sup>16</sup> A similar calculation can be made for E2.<sup>17</sup> Only free testosterone is biologically active, and it best indicates the clinical situation of the patient. Free and bioavailable testosterone are also referred to as non-SHBG-bound testosterone and can be obtained by precipitation of the SHBG-bound-testosterone with ammonium sulfate, and by equilibrium dialysis.<sup>18</sup>

Elevated SHBG levels can be seen in elderly men, and are often found in patients with hyperthyroidism and cirrhosis. SHBG levels also increase when oral contraceptives or antiepileptic drugs are taken. Pregnant women have markedly higher SHBG serum concentrations due to their increased estrogen production. Decreased SHBG concentrations are often seen in patients with hypothyroidism, polycystic ovarian syndrome, obesity, hirsutism, elevated androgen levels, alopecia, and acromegaly. The Elecsys SHBG assay employs two monoclonal antibodies specifically directed against human SHBG.

Cross-reactivity with  $\alpha_1$ -fetoprotein (AFP), corticosteroid binding globulin (CBG), DHT, E2, fibrinogen, human immunoglobulin A (IgA), human immunoglobulin G (IgG), plasminogen, thyroxine binding globulin (TBG), testosterone, thyroglobulin (Tg), transferrin, and thyrotropin (TSH) is negligible.

## Test principle

Sandwich principle. Total duration of assay: 18 minutes.

- 1st incubation: 10  $\mu$ L of sample, a biotinylated monoclonal SHBG-specific antibody, and a monoclonal SHBG-specific antibody labeled with a ruthenium complex<sup>a)</sup> form a sandwich complex.
- 2nd incubation: After addition of streptavidin-coated microparticles, the complex becomes bound to the solid phase via interaction of biotin and streptavidin.
- The reaction mixture is aspirated into the measuring cell where the microparticles are magnetically captured onto the surface of the electrode. Unbound substances are then removed with ProCell/ProCell M. Application of a voltage to the electrode then induces chemiluminescent emission which is measured by a photomultiplier.
- Results are determined via a calibration curve which is instrument-specifically generated by 2-point calibration and a master curve provided via the reagent barcode or e-barcode.

a) Tris(2,2'-bipyridyl)ruthenium(II)-complex (Ru(bpy)<sub>3</sub><sup>2+</sup>)

## Reagents - working solutions

The reagent rackpack is labeled as SHBG.

- M Streptavidin-coated microparticles (transparent cap), 1 bottle, 6.5 mL:  
Streptavidin-coated microparticles 0.72 mg/mL; preservative.
- R1 Anti-SHBG-Ab~biotin (gray cap), 1 bottle, 10 mL:  
Biotinylated monoclonal anti-SHBG antibody (mouse) 1.25 mg/L;  
phosphate buffer 100 mmol/L, pH 7.2; preservative.
- R2 Anti-SHBG-Ab~Ru(bpy)<sub>3</sub><sup>2+</sup> (black cap), 1 bottle, 10 mL:  
Monoclonal anti-SHBG antibody (mouse) labeled with ruthenium complex 1.25 mg/L; phosphate buffer 100 mmol/L, pH 7.2; preservative.

## Precautions and warnings

For in vitro diagnostic use for health care professionals. Exercise the normal precautions required for handling all laboratory reagents.

Infectious or microbial waste:

Warning: handle waste as potentially biohazardous material. Dispose of waste according to accepted laboratory instructions and procedures.

Environmental hazards:

Apply all relevant local disposal regulations to determine the safe disposal.

Safety data sheet available for professional user on request.

This kit contains components classified as follows in accordance with the Regulation (EC) No. 1272/2008:



## Warning

H317 May cause an allergic skin reaction.

## Prevention:

P261 Avoid breathing dust/fume/gas/mist/vapours/spray.

P272 Contaminated work clothing should not be allowed out of the workplace.

P280 Wear protective gloves.

## Response:

P333 + P313 If skin irritation or rash occurs: Get medical advice/attention.

P362 + P364 Take off contaminated clothing and wash it before reuse.

## Disposal:

P501 Dispose of contents/container to an approved waste disposal plant.

Product safety labeling follows EU GHS guidance.

Contact phone: all countries: +49-621-7590

Avoid foam formation in all reagents and sample types (specimens, calibrators and controls).

## Reagent handling

The reagents in the kit have been assembled into a ready-for-use unit that cannot be separated.

All information required for correct operation is read in from the respective reagent barcodes.

## Storage and stability

Store at 2-8 °C.

Do not freeze.

Store the Elecsys reagent kit **upright** in order to ensure complete availability of the microparticles during automatic mixing prior to use.

Stability:	
unopened at 2-8 °C	up to the stated expiration date
after opening at 2-8 °C	12 weeks
on the analyzers	7 weeks

## Specimen collection and preparation

Only the specimens listed below were tested and found acceptable.

Serum collected using standard sampling tubes or tubes containing separating gel, or lithium heparin plasma.

Do not use EDTA plasma.

Criterion: Recovery within 90-110 % of serum value or slope 0.9-1.1 + intercept within  $\pm 2 \times$  analytical sensitivity (LDL) + coefficient of correlation  $> 0.95$ .

Stable for 3 days at 2-8 °C, 1 month at -20 °C ( $\pm 5$  °C). Freeze only once.<sup>19</sup>

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube manufacturer.

Centrifuge samples containing precipitates before performing the assay.

Do not use heat-inactivated samples.

Do not use samples and controls stabilized with azide.

Ensure the samples, calibrators and controls are at 20-25 °C prior to measurement.

Due to possible evaporation effects, samples, calibrators and controls on the analyzers should be analyzed/measured within 2 hours.

## Materials provided

See "Reagents – working solutions" section for reagents.

## Materials required (but not provided)

- [REF] 03052028190, SHBG CalSet, for 4 x 1.0 mL
- [REF] 11731416190, PreciControl Universal, for 4 x 3.0 mL
- [REF] 03609987190, Diluent MultiAssay, 2 x 16 mL sample diluent

- General laboratory equipment

- **cobas e** analyzer

Additional materials for the **cobas e** 411 analyzer:

- [REF] 11662988122, ProCell, 6 x 380 mL system buffer
- [REF] 11662970122, CleanCell, 6 x 380 mL measuring cell cleaning solution
- [REF] 11930346122, Elecsys SysWash, 1 x 500 mL washwater additive
- [REF] 11933159001, Adapter for SysClean
- [REF] 11706802001, AssayCup, 60 x 60 reaction cups
- [REF] 11706799001, AssayTip, 30 x 120 pipette tips
- [REF] 11800507001, Clean-Liner

Additional materials for **cobas e** 601 and **cobas e** 602 analyzers:

- [REF] 04880340190, ProCell M, 2 x 2 L system buffer
- [REF] 04880293190, CleanCell M, 2 x 2 L measuring cell cleaning solution
- [REF] 03023141001, PC/CC-Cups, 12 cups to prewarm ProCell M and CleanCell M before use
- [REF] 03005712190, ProbeWash M, 12 x 70 mL cleaning solution for run finalization and rinsing during reagent change
- [REF] 12102137001, AssayTip/AssayCup, 48 magazines x 84 reaction cups or pipette tips, waste bags
- [REF] 03023150001, WasteLiner, waste bags
- [REF] 03027651001, SysClean Adapter M

Additional materials for all analyzers:

- [REF] 11298500316, ISE Cleaning Solution/Elecsys SysClean, 5 x 100 mL system cleaning solution

## Assay

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

Resuspension of the microparticles takes place automatically prior to use. Read in the test-specific parameters via the reagent barcode. If in exceptional cases the barcode cannot be read, enter the 15-digit sequence of numbers.

Bring the cooled reagents to approximately 20 °C and place on the reagent disk (20 °C) of the analyzer. Avoid foam formation. The system automatically regulates the temperature of the reagents and the opening/closing of the bottles.

## Calibration

Traceability: This method has been standardized against the 1st International Standard for SHBG from the National Institute for Biological Standards and Control (NIBSC) code 95/560.

Every Elecsys reagent set has a barcoded label containing specific information for calibration of the particular reagent lot. The predefined master curve is adapted to the analyzer using the relevant CalSet.

**Calibration frequency:** Calibration must be performed once per reagent lot using fresh reagent (i.e. not more than 24 hours since the reagent kit was registered on the analyzer).

Calibration interval may be extended based on acceptable verification of calibration by the laboratory.

Renewed calibration is recommended as follows:

- after 1 month (28 days) when using the same reagent lot
- after 7 days (when using the same reagent kit on the analyzer)
- as required: e.g. quality control findings outside the defined limits

## Quality control

For quality control, use PreciControl Universal.

In addition, other suitable control material can be used.

Controls for the various concentration ranges should be run individually at least once every 24 hours when the test is in use, once per reagent kit, and following each calibration.

The control intervals and limits should be adapted to each laboratory's individual requirements. Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

If necessary, repeat the measurement of the samples concerned.

Follow the applicable government regulations and local guidelines for quality control.

## Calculation

The analyzer automatically calculates the analyte concentration of each sample either in nmol/L, µg/mL or mg/L (selectable).

Conversion factors: nmol/L x 0.095 = µg/mL (mg/L)  
µg/mL (mg/L) x 10.53 = nmol/L

## Limitations - interference

The assay is unaffected by icterus (bilirubin < 1026 µmol/L or < 60 mg/dL), hemolysis (Hb < 1.8 mmol/L or < 2.9 g/dL), lipemia (Intralipid < 2700 mg/dL) and biotin (< 246 nmol/L or < 60 ng/mL).

Criterion: Recovery within ± 10 % of initial value.

Samples should not be taken from patients receiving therapy with high biotin doses (i.e. > 5 mg/day) until at least 8 hours following the last biotin administration.

No interference was observed from rheumatoid factors up to a concentration of 1160 IU/mL.

There is no high-dose hook effect at SHBG concentrations up to 1000 nmol/L.

In vitro tests were performed on 16 commonly used pharmaceuticals. No interference with the assay was found.

In rare cases, interference due to extremely high titers of antibodies to analyte-specific antibodies, streptavidin or ruthenium can occur. These effects are minimized by suitable test design.

For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.

## Limits and ranges

### Measuring range

0.350-200 nmol/L (defined by the lower detection limit and the maximum of the master curve). Values below the lower detection limit are reported as < 0.350 nmol/L. Values above the measuring range are reported as > 200 nmol/L (or up to 2000 nmol/L for 10-fold diluted samples).

### Lower limits of measurement

#### Lower detection limit of the test

Lower detection limit: 0.350 nmol/L

The Lower Detection Limit represents the lowest measurable analyte level that can be distinguished from zero. It is calculated as the value lying two standard deviations above that of the lowest standard (master calibrator, standard 1 + 2 SD, repeatability study, n = 21).

### Dilution

Samples with SHBG concentrations above the measuring range can be diluted with Diluent MultiAssay. The recommended dilution is 1:10 (either automatically by the analyzers, or manually). The concentration of the diluted sample must be > 20 nmol/L.

After manual dilution, multiply the result by the dilution factor.

After dilution by the analyzers, the software automatically takes the dilution into account when calculating the sample concentration.

### Expected values

The following table shows the results obtained from a group of 415 males and 343 females using the Elecsys SHBG assay. All subjects were apparently healthy, non-obese (BMI, body mass index ≤ 30), non-pregnant adults without intake of any contraceptive or relevant prescription drugs (study number CIM 000669). Blood samples (fasting) were taken between 6.30 am and 2.00 pm. This clinical study focusing on the Elecsys Testosterone II assay included measurements in parallel using the Elecsys SHBG assay. Please refer to the Elecsys Testosterone II package insert for SHBG values in combination with testosterone.

## SHBG

	SHBG (nmol/L)		
	N	Median	5-95 <sup>th</sup> percentiles
Males (20-49 years)	241	33.2	18.3-54.1
Males (≥ 50 years)	174	40.6	20.6-76.7
Females (20-49 years)	166	67.8	32.4-128
Females (≥ 50 years)	177	62.4	27.1-128

Each laboratory should investigate the transferability of the expected values to its own patient population and if necessary determine its own reference ranges.

## Specific performance data

Representative performance data on the analyzers are given below. Results obtained in individual laboratories may differ.

### Precision

Precision was determined using Elecsys reagents, pooled human sera and controls in a modified protocol (EP5-A) of the CLSI (Clinical and Laboratory Standards Institute): 6 times daily for 10 days (n = 60); repeatability on MODULAR ANALYTICS E170 analyzer, n = 21. The following results were obtained:

cobas e 411 analyzer					
		Repeatability		Intermediate precision	
Sample	Mean nmol/L	SD nmol/L	CV %	SD nmol/L	CV %
Human serum 1	14.1	0.30	2.1	0.39	2.7
Human serum 2	44.2	1.05	2.4	1.24	2.8
Human serum 3	204	5.61	2.7	11.4	5.6
PreciControl U <sup>b)</sup> 1	34.9	0.76	2.2	0.92	2.6
PreciControl U2	19.0	0.41	2.2	0.52	2.7

b) U = Universal

cobas e 601 and cobas e 602 analyzers						
		Repeatability			Intermediate precision	
Sample	Mean nmol/L	SD nmol/L	CV %	Mean nmol/L	SD nmol/L	CV %
Human serum 1	14.9	0.17	1.1	13.7	0.24	1.8
Human serum 2	45.7	0.60	1.3	42.0	0.89	2.1
Human serum 3	219	3.76	1.7	189	7.58	4.0
PreciControl U1	35.3	0.46	1.3	33.1	0.63	1.9
PreciControl U2	19.2	0.21	1.1	18.1	0.42	2.3

## Method comparison

A comparison of the Elecsys SHBG assay (y) with a commercially available SHBG test (x) using clinical samples gave the following correlations:

Number of samples measured: 109

Passing/Bablok<sup>20</sup> Linear regression  
 $y = 1.17x - 3.26$   $y = 1.15x - 1.82$   
 $r = 0.909$   $r = 0.981$

The sample concentrations were between 11.2 and 155 nmol/L.

## Analytical specificity

For the monoclonal antibodies used, non-detectable cross-reactivities were found for the following substances:

AFP, CBG, DHT, estradiol, fibrinogen, human IgA, human IgG, plasminogen, TBG, testosterone, Tg, transferrin, and TSH.

## References

- Petra PH. The plasma sex steroid binding protein (SBP or SHBG). A critical review of recent developments on the structure, molecular biology and function. J Steroid Biochem Molec Biol 1991;40:735-753.
- Avvakumov GV, Grishkovskaya I, Muller IA, et al. Resolution of the human sex hormone-binding globulin dimer interface and evidence for two steroid-binding sites per homodimer. J Biol Chem 2001;276:34453-34457.
- Hammond GL, Bochinfuso WP. Sex hormone-binding globulin/androgen-binding protein: steroid-binding and dimerization domains. J Steroid Biochem Molec Biol 1995;53:543-552.
- Davis S. Testosterone deficiency in women. J Reprod Med 2001;46:291-306.
- Burger HG. Androgen production in women. Fertil Steril 2002;77:3-5.
- Simó R, Sáez-López C, Barbosa-Desongles A, et al. Novel insights in SHBG regulation and clinical implications. Trends in Endocrinol Metab 2015;26(7):376-383.
- Bonnet F, Velayoudom Cephiset FL, Gautier A, et al. Role of sex steroids, intrahepatic fat and liver enzymes in the association between SHBG and metabolic features. Clin Endocrinol 2013;79:517-522.
- Kavanagh K, Espeland MA, Sutton-Tyrrell K, et al. Liver Fat and SHBG Affect insulin Resistance in Midlife Women: The Study of Women's Health Across the Nation (SWAN). Obesity 2013;21:1031-1038.
- Maggio M, Ceda GP, Lauretani F, et al. SHBG, Sex Hormones, and Inflammatory Markers in Older Women. J Clin Endocrinol Metab 2011;96(4):1053-1059.
- Maggio M, Cattabiani C, Lauretani F, et al. SHBG and endothelial function in older subject. Int J of Cardiology 2013;168(3):2825-2830.
- Calderon-Margalit R, Schwartz S.M, Wellons MF. Prospective Association of Serum Androgens and Sex Hormone-Binding Globulin with Subclinical Cardiovascular Disease in Young Adult Women: The "Coronary Artery Risk Development in Young Adults" Women's Study. J Clin Endocrinol Metab 2010;95(9):4424-4431.
- Hedderon MM, Xu F, Darbinian JA, et al. Prepregnancy SHBG Concentrations and Risk for Subsequently Developing Gestational Diabetes Mellitus. Diabetes Care 2014;37(5):1296-1303.
- Soriguer F, Rubio-Martín E, Fernández D, et al. Testosterone, SHBG and risk of type 2 diabetes in the second evaluation of the Pizarra cohort study. Eur J Clin Invest 2012;42(1):79-85.
- Fortunati N, Catalano MG, Boccuzzi G, et al. Sex Hormone-Binding (SHBG), estradiol and breast cancer. J Mol Cell Endocrinol 2010;316(1):86-92.
- Pugeat M, Crave JC, Tourniaire J, et al. Clinical utility of sex hormone-binding globulin measurement. Horm Res 1996;45:148-155.
- Zakharov MN, Bhasin S, Travison TG, et al. A multi-step, dynamic allosteric model of testosterone's binding to sex hormone binding globulin. Mol Cell Endocrinol 2015;399:190-200.
- Rosner W. Free estradiol and sex hormone-binding globulin. Steroids 2015;99:113-116.
- Morley JE, Patrick P, Perry III HM. Evaluation of assays available to measure free testosterone. Metabolism 2002;51:554-609.
- Wu AHB. Tietz Clinical Guide To Laboratory Tests, 4th Edition, WB Saunders Co, 2006: 982 pp.
- Bablok W, Passing H, Bender R, et al. A general regression procedure for method transformation. Application of linear regression procedures for method comparison studies in clinical chemistry, Part III. J Clin Chem Clin Biochem 1988 Nov;26(11):783-790.

For further information, please refer to the appropriate operator's manual for the analyzer concerned, the respective application sheets and the Method Sheets of all necessary components (if available in your country).

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the Member State in which the user and/or the patient is established.

## Symbols

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard (for USA: see [dialog. Roche.com](http://dialog. Roche.com) for definition of symbols used):

	Contents of kit
	Analyzers/Instruments on which reagents can be used
	Reagent
	Calibrator
	Volume for reconstitution
	Global Trade Item Number

COBAS, COBAS E, ELECSYS and PRECICONTROL are trademarks of Roche. INTRALIPID is a trademark of Fresenius Kabi AB.

All other product names and trademarks are the property of their respective owners.

Additions, deletions or changes are indicated by a change bar in the margin.

© 2022, Roche Diagnostics



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim  
[www.Roche.com](http://www.Roche.com)

+800 5505 6606

