

## Urea/BUN

### Order information

REF	CONTENT	Analyzer(s) on which <b>cobas c</b> pack(s) can be used
08058806190	Urea/BUN (600 tests)	System-ID 2119 001 <b>cobas c 303, cobas c 503</b>
Materials required (but not provided):		
10759350190	Calibrator f.a.s. (12 x 3 mL)	Code 20401
05117003190	PreciControl ClinChem Multi 1 (20 x 5 mL)	Code 20391
05947626190	PreciControl ClinChem Multi 1 (4 x 5 mL)	Code 20391
05117216190	PreciControl ClinChem Multi 2 (20 x 5 mL)	Code 20392
05947774190	PreciControl ClinChem Multi 2 (4 x 5 mL)	Code 20392
08063494190	Diluent NaCl 9 % (123 mL)	System-ID 2906 001

### English

#### System information

**UREAL:** ACN 21191 (Serum/plasma)

**URELU:** ACN 21190 (Urine)

**U-BUN:** ACN 21192 (Serum/plasma)

**UBUNU:** ACN 21193 (Urine)

#### Intended use

In vitro test for the quantitative determination of urea/urea nitrogen in human serum, plasma and urine on Roche/Hitachi **cobas c** systems.

#### Summary<sup>1</sup>

Urea is the major end product of protein nitrogen metabolism. It is synthesized by the urea cycle in the liver from ammonia which is produced by amino acid deamination. Urea is excreted mostly by the kidneys but minimal amounts are also excreted in sweat and degraded in the intestines by bacterial action.


Determination of blood urea nitrogen is the most widely used screening test for renal function. When used in conjunction with serum creatinine determinations it can aid in the differential diagnosis of the three types of azotemia: prerenal, renal and postrenal.

Elevations in blood urea nitrogen concentration are seen in inadequate renal perfusion, shock, diminished blood volume (prerenal causes), chronic nephritis, nephrosclerosis, tubular necrosis, glomerular nephritis (renal causes) and urinary tract obstruction (postrenal causes). Transient elevations may also be seen during periods of high protein intake. Unpredictable levels occur with liver diseases.


#### Test principle

Kinetic test with urease and glutamate dehydrogenase.<sup>2,3,4,5</sup>

Urea is hydrolyzed by urease to form ammonium and carbonate.



In the second reaction 2-oxoglutarate reacts with ammonium in the presence of glutamate dehydrogenase (GLDH) and the coenzyme NADH to produce L-glutamate. In this reaction two moles of NADH are oxidized to NAD<sup>+</sup> for each mole of urea hydrolyzed.



The rate of decrease in the NADH concentration is directly proportional to the urea concentration in the specimen and is measured photometrically.

#### Reagents - working solutions

**R1** NaCl 9 %

**R3** TRIS buffer: 220 mmol/L, pH 8.6; 2-oxoglutarate: 73 mmol/L; NADH: 2.5 mmol/L; ADP: 6.5 mmol/L; urease (jack bean): ≥ 300 μkat/L; GLDH (bovine liver): ≥ 80 μkat/L; preservative; nonreactive stabilizers

R1 is in position B and R3 is in position C.

#### Precautions and warnings

For in vitro diagnostic use for health care professionals. Exercise the normal precautions required for handling all laboratory reagents.

Infectious or microbial waste:

Warning: handle waste as potentially biohazardous material. Dispose of waste according to accepted laboratory instructions and procedures.

Environmental hazards:

Apply all relevant local disposal regulations to determine the safe disposal.

Safety data sheet available for professional user on request.

#### Reagent handling

Ready for use

#### Storage and stability

Shelf life at 2-8 °C:

See expiration date on **cobas c** pack label.

On-board in use and refrigerated on the analyzer:

8 weeks

#### Specimen collection and preparation

For specimen collection and preparation only use suitable tubes or collection containers.

Only the specimens listed below were tested and found acceptable.

Serum

Plasma: Li-heparin and K<sub>2</sub>-EDTA plasma. Do not use ammonium heparin.

The sample types listed were tested with a selection of sample collection tubes that were commercially available at the time of testing, i.e. not all available tubes of all manufacturers were tested. Sample collection systems from various manufacturers may contain differing materials which could affect the test results in some cases. When processing samples in primary tubes (sample collection systems), follow the instructions of the tube manufacturer.

Urine

Bacterial growth in the specimen and high atmospheric ammonia concentrations as well as contamination by ammonium ions may cause erroneously elevated results. If stabilizers are added to the sample, the sample index feature must not be used.

Stability in *serum/plasma*.<sup>6</sup>

7 days at 15-25 °C

7 days at 2-8 °C

1 year at (-15)-(-25) °C

Stability in *urine*.<sup>6</sup>

2 days at 15-25 °C

7 days at 2-8 °C

1 month at (-15)-(-25) °C

Centrifuge samples containing precipitates before performing the assay.

See the limitations and interferences section for details about possible sample interferences.

#### Materials provided

See "Reagents – working solutions" section for reagents.

#### Materials required (but not provided)

See "Order information" section

General laboratory equipment

**Assay**

For optimum performance of the assay follow the directions given in this document for the analyzer concerned. Refer to the appropriate operator's manual for analyzer-specific assay instructions.

The performance of applications not validated by Roche is not warranted and must be defined by the user.

**Application for serum and plasma****Test definition**

Reporting time	10 min		
Wavelength (sub/main)	700/340 nm		
Reagent pipetting		Diluent (H <sub>2</sub> O)	
R1	8 µL	66 µL	
R3	28 µL	81 µL	
Sample volumes	Sample	Sample dilution	
		Sample	Diluent (NaCl)
Normal	1.5 µL	–	–
Decreased	1.5 µL	25 µL	50 µL
Increased	1.5 µL	–	–

**Application for urine****Test definition**

Reporting time	10 min		
Wavelength (sub/main)	700/340 nm		
Reagent pipetting		Diluent (H <sub>2</sub> O)	
R1	8 µL	66 µL	
R3	28 µL	81 µL	
Sample volumes	Sample	Sample dilution	
		Sample	Diluent (NaCl)
Normal	1.5 µL	2.0 µL	98 µL
Decreased	1.5 µL	1.3 µL	116 µL
Increased	1.5 µL	–	–

For further information about the assay test definitions refer to the application parameters setting screen of the corresponding analyzer and assay.

**Calibration***Application for serum/plasma (ACN 21191/21192)*

Calibrators	S1: H <sub>2</sub> O
	S2: C.f.a.s.
Calibration mode	Linear
Calibration frequency	Full calibration
	- after reagent lot change
	- after 4 weeks on-board
	- as required following quality control procedures

*Application for urine (ACN 21190/21193)*

Transfer of calibration from serum/plasma application (ACN 21191/21192)

Calibration interval may be extended based on acceptable verification of calibration by the laboratory.

Traceability: This method has been standardized against ID/MS.

**Quality control**

For quality control, use control materials as listed in the "Order information" section. In addition, other suitable control material can be used.

Serum/plasma:	PreciControl ClinChem Multi 1, PreciControl ClinChem Multi 2
---------------	--

Urine: Quantitative urine controls are recommended for routine quality control.

The control intervals and limits should be adapted to each laboratory's individual requirements. It is recommended to perform quality control always after lot calibration and subsequently at least every 8 weeks.

Values obtained should fall within the defined limits. Each laboratory should establish corrective measures to be taken if values fall outside the defined limits.

Follow the applicable government regulations and local guidelines for quality control.

**Calculation**

**cobas c** systems automatically calculate the analyte concentration of each sample in the unit mmol/L (mg/dL, g/L).

Conversion factors:

mmol/L urea x 6.006 = mg/dL urea

mmol/L urea x 0.06006 = g/L urea

mmol/L urea nitrogen x 2.801 = mg/dL urea nitrogen

mmol/L urea nitrogen x 0.02801 = g/L urea nitrogen

When 24-hour urine is used as the specimen, multiply the result by the 24-hour volume to obtain values in g or mmol/24 hours.

**Limitations - interference**

Criterion: Recovery within ± 10 % of initial value at an urea concentration of 8.3 mmol/L (49.8 mg/dL urea, 23.2 mg/dL urea nitrogen) in serum/plasma and at an urea concentration of 150 mmol/L (901 mg/dL urea, 421 mg/dL urea nitrogen) in urine. Recovery within ± 10 % for drug interference.

*Serum/plasma*

Icterus:<sup>7</sup> No significant interference up to an I index of 60 for conjugated and unconjugated bilirubin (approximate conjugated and unconjugated bilirubin concentration: 1026 µmol/L or 60 mg/dL).

Hemolysis:<sup>7</sup> No significant interference up to an H index of 1000 (approximate hemoglobin concentration: 621 µmol/L or 1000 mg/dL).

Lipemia (Intralipid):<sup>7</sup> No significant interference up to an L index of 1000. There is poor correlation between the L index (corresponds to turbidity) and triglycerides concentration.

Ammonium ions may cause erroneously elevated results.

Drugs: No interference was found at therapeutic concentrations using common drug panels.<sup>8,9</sup>

In very rare cases, gammopathy, in particular type IgM (Waldenström's macroglobulinemia), may cause unreliable results.<sup>10</sup>

*Urine*

Hemolysis: No significant interference up to an H index of 750 (approximate hemoglobin concentration: 466 µmol/L or 750 mg/dL).

Drugs: No interference was found at therapeutic concentrations using common drug panels.<sup>9</sup>

For diagnostic purposes, the results should always be assessed in conjunction with the patient's medical history, clinical examination and other findings.

**ACTION REQUIRED**

**Special Wash Programming:** The use of special wash steps is mandatory when certain test combinations are run together on **cobas c** systems. All special wash programming necessary for avoiding carry-over is available via the **cobas** link. The latest version of the carry-over evasion list can be found with the NaOHD/SMS/SCCS Method Sheet for information. For further instructions refer to the operator's manual.

**Limits and ranges****Measuring range***Serum/plasma*

0.5-40 mmol/L (3.0-240 mg/dL urea, 1.4-112 mg/dL urea nitrogen)

Determine samples having higher concentrations via the rerun function. Dilution of samples via the rerun function is a 1:3 dilution. Results from samples diluted using the rerun function are automatically multiplied by a factor of 3.

*Urine*

1-2000 mmol/L (6-12000 mg/dL urea, 2.8-5600 mg/dL urea nitrogen)

Determine samples having higher concentrations via the rerun function. Dilution of samples via the rerun function is a 1:1.8 dilution. Results from samples diluted using the rerun function are automatically multiplied by a factor of 1.8.

Determine samples having concentrations lower than the technical limit of 40 mmol/L (240 mg/dL urea and 112 mg/dL urea nitrogen) via the rerun function. Samples are measured undiluted.

**Lower limits of measurement***Limit of Blank, Limit of Detection and Limit of Quantitation**Serum/plasma*

Limit of Blank = 0.5 mmol/L

Limit of Detection = 0.5 mmol/L

Limit of Quantitation = 0.5 mmol/L

*Urine*

Limit of Blank = 1.0 mmol/L

Limit of Detection = 1.0 mmol/L

Limit of Quantitation = 1.0 mmol/L

The Limit of Blank, Limit of Detection and Limit of Quantitation were determined in accordance with the CLSI (Clinical and Laboratory Standards Institute) EP17-A2 requirements.

The Limit of Blank is the 95<sup>th</sup> percentile value from  $n \geq 60$  measurements of analyte-free samples over several independent series. The Limit of Blank corresponds to the concentration below which analyte-free samples are found with a probability of 95 %.

The Limit of Detection is determined based on the Limit of Blank and the standard deviation of low concentration samples.

The Limit of Detection corresponds to the lowest analyte concentration which can be detected (value above the Limit of Blank with a probability of 95 %).

The Limit of Quantitation is the lowest analyte concentration that can be reproducibly measured with a total error of 20 %. It has been determined using low concentration urea/urea nitrogen samples.

**Expected values****mmol/L**

Urea:

*Serum/plasma*<sup>11</sup>

Adults 2.76-8.07 mmol/L

*Urine*

24-hour urine<sup>12</sup> 428-714 mmol/24 h,  
corresponding to  
286-595 mmol/L<sup>a)</sup>

a) Based on average urine output of 1.2-1.5 L/24 h

Urea nitrogen (BUN):

*Serum/plasma*<sup>12</sup>

Adults (18-60 years) 2.14-7.14 mmol/L

Adults (60-90 years) 2.86-8.21 mmol/L

Infants (< 1 year) 1.43-6.78 mmol/L

Infants/children 1.79-6.43 mmol/L

*Urine*

24-hour urine<sup>12</sup> 428-714 mmol/24 h,  
corresponding to  
286-595 mmol/L<sup>b)</sup>

b) Based on average urine output of 1.2-1.5 L/24 h

**mg/dL**

Urea:

*Serum/plasma*<sup>11</sup>

Adults 16.6-48.5 mg/dL

*Urine*

24-hour urine<sup>12</sup> 25.7-42.9 g/24 h,  
corresponding to  
1.71-3.57 g/dL<sup>a)</sup>

a) Based on average urine output of 1.2-1.5 L/24 h

Urea nitrogen (BUN):

*Serum/plasma*<sup>12</sup>

Adults (18-60 years) 6-20 mg/dL

Adults (60-90 years) 8-23 mg/dL

Infants (< 1 year) 4-19 mg/dL

Infants/children 5-18 mg/dL

*Urine*

24-hour urine<sup>12</sup> 12-20 g/24 h,  
corresponding to  
801-1666 mg/dL<sup>b)</sup>

b) Based on average urine output of 1.2-1.5 L/24 h

Each laboratory should investigate the transferability of the expected values to its own patient population and if necessary determine its own reference ranges.

**Specific performance data**

Representative performance data on the analyzers are given below. These data represent the performance of the analytical procedure itself.

Results obtained in individual laboratories may differ due to heterogeneous sample materials, aging of analyzer components and mixture of reagents running on the analyzer.

**Precision**

Precision was determined using human samples and controls in accordance with the CLSI (Clinical and Laboratory Standards Institute) EP05-A3 requirements with repeatability ( $n = 84$ ) and intermediate precision (2 aliquots per run, 2 runs per day, 21 days). Results for repeatability and intermediate precision were obtained on the **cobas c 503** analyzer.

*Serum/plasma*

*Repeatability*

	Mean	SD	CV
	mmol/L	mmol/L	%
PCCC1 <sup>c)</sup>	6.53	0.0408	0.6
PCCC2 <sup>d)</sup>	18.3	0.0690	0.4
Human serum 1	1.31	0.0416	3.2
Human serum 2	5.12	0.0441	0.9
Human serum 3	7.67	0.0451	0.6
Human serum 4	18.7	0.101	0.5
Human serum 5	31.0	0.124	0.4

*Urine*

Repeatability	Mean	SD	CV
	mmol/L	mmol/L	%
Control 1 <sup>e)</sup>	143	2.86	2.0
Control 2 <sup>e)</sup>	239	3.68	1.5
Human urine 1	3.22	0.0435	1.4
Human urine 2	73.2	2.50	3.4
Human urine 3	407	3.28	0.8
Human urine 4	922	5.03	0.5
Human urine 5	1583	10.1	0.6

Intermediate precision	Mean	SD	CV
	mmol/L	mmol/L	%
Control 1 <sup>e)</sup>	143	3.17	2.2
Control 2 <sup>e)</sup>	239	4.32	1.8
Human urine 1	3.22	0.0547	1.7
Human urine 2	73.2	2.78	3.8
Human urine 3	411	4.93	1.2
Human urine 4	919	11.5	1.2
Human urine 5	1583	19.7	1.2

e) commercially available control material

The data obtained on **cobas c 503** analyzer(s) are representative for **cobas c 303** analyzer(s).

**Method comparison**

Urea values for human serum, plasma and urine samples obtained on a **cobas c 503** analyzer (y) were compared with those determined using the corresponding reagent on a **cobas c 501** analyzer (x).

*Serum/plasma*

Sample size (n) = 94

Passing/Bablok <sup>13</sup>	Linear regression
$y = 1.009x + 0.0202$ mmol/L	$y = 1.006x + 0.0265$ mmol/L
$\tau = 0.986$	$r = 1.000$

The sample concentrations were between 0.600 and 38.1 mmol/L.

*Urine*

Sample size (n) = 91

Passing/Bablok <sup>13</sup>	Linear regression
$y = 0.962x - 0.432$ mmol/L	$y = 0.960x + 0.586$ mmol/L
$\tau = 0.982$	$r = 1.000$

The sample concentrations were between 71.0 and 1964 mmol/L.

Urea values for human serum, plasma and urine samples obtained on a **cobas c 303** analyzer (y) were compared with those determined using the corresponding reagent on a **cobas c 501** analyzer (x).

*Serum/plasma*

Sample size (n) = 89

Passing/Bablok <sup>13</sup>	Linear regression
$y = 1.017x + 0.0905$ mmol/L	$y = 1.015x + 0.148$ mmol/L
$\tau = 0.986$	$r = 1.000$

The sample concentrations were between 0.700 and 35.4 mmol/L.

*Urine*

Sample size (n) = 73

Passing/Bablok <sup>13</sup>	Linear regression
$y = 0.981x + 0.901$ mmol/L	$y = 0.973x + 4.74$ mmol/L
$\tau = 0.960$	$r = 0.999$

The sample concentrations were between 41.0 and 1875 mmol/L.

**References**




- Rock RC, Walker WG, Jennings CD. Nitrogen metabolites and renal function. In: Tietz NW, ed. *Fundamentals of Clinical Chemistry*. 3rd ed. Philadelphia: WB Saunders 1987;669–704.
- Richterich R, Colombo JP. *Klinische Chemie*. 4th ed. Basel: Karger S 1978;319-324.
- Talke H, Schubert GA. Enzymatische Harnstoffbestimmung in Blut und Serum im optischen Test nach Warburg. *Klin Wochenschr* 1965;43:174.
- Tiffany TO, Jansen JM, Burtis CA, et al. Enzymatic kinetic rate and end-point analyses of substrate, by use of a GeMSAEC Fast Analyzer. *Clin Chem* 1972;18:829-840.
- Sampson EJ, Baired MA, Burtis CA, et al. A coupled-enzyme equilibrium method for measuring urea in serum: Optimization and evaluation of the AACC study group on urea candidate reference method. *Clin Chem* 1980;26:816-826.
- WHO Publication: Use of anticoagulants in diagnostic laboratory investigations, WHO/DIL/LAB/99.1 Rev.2:Jan 2002.
- Glick MR, Ryder KW, Jackson SA. Graphical Comparisons of Interferences in Clinical Chemistry Instrumentation. *Clin Chem* 1986;32:470-475.
- Breuer J. Report on the Symposium "Drug effects in Clinical Chemistry Methods". *Eur J Clin Chem Clin Biochem* 1996;34:385-386.
- Sonntag O, Scholer A. Drug interference in clinical chemistry: recommendation of drugs and their concentrations to be used in drug interference studies. *Ann Clin Biochem* 2001;38:376-385.
- Bakker AJ, Mücke M. Gammopathy interference in clinical chemistry assays: mechanisms, detection and prevention. *Clin Chem Lab Med* 2007;45(9):1240-1243.
- Löhr B, El-Samalouti V, Junge W, et al. Reference Range Study for Various Parameters on Roche Clinical Chemistry Analyzers. *Clin Lab* 2009;55:465-471.
- Wu AHB, ed. *Tietz Clinical Guide to Laboratory Tests*, 4th edition. St. Louis (MO): Saunders Elsevier 2006;1096.
- Bablok W, Passing H, Bender R, et al. A general regression procedure for method transformation. Application of linear regression procedures for method comparison studies in clinical chemistry, Part III. *J Clin Chem Clin Biochem* 1988 Nov;26(11):783-790.

A point (period/stop) is always used in this Method Sheet as the decimal separator to mark the border between the integral and the fractional parts of a decimal numeral. Separators for thousands are not used.

Any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the Member State in which the user and/or the patient is established.

**Symbols**

Roche Diagnostics uses the following symbols and signs in addition to those listed in the ISO 15223-1 standard (for USA: see [dialog.roche.com](http://dialog.roche.com) for definition of symbols used):

	Contents of kit
	Volume after reconstitution or mixing
	Global Trade Item Number

COBAS, COBAS C and PRECICONTROL are trademarks of Roche.

All other product names and trademarks are the property of their respective owners.

Additions, deletions or changes are indicated by a change bar in the margin.

© 2021, Roche Diagnostics

0108058806190c503V5.0

# UREAL

Urea/BUN

CE 0123



Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim  
[www.roche.com](http://www.roche.com)

+800 5505 6606



# cobas®